



Certificate of Analysis

COMPLIANCE FOR RETAIL

Sample: DA30331004-007
Harvest/Lot ID: 0695 6433 5239 0312
Batch#: 0695 6433 5239 0312
Cultivation Facility: Tampa Cultivation
Processing Facility : Tampa Processing
Source Facility : Tampa Cultivation
Seed to Sale# 7688 3965 4841 3841
Batch Date: 03/16/23
Sample Size Received: 26 gram
Total Amount: 1132 units
Retail Product Size: 1 gram
Ordered : 03/30/23
Sampled : 03/30/23
Completed: 04/03/23
Sampling Method: SOP.T.20.010
Apr 03, 2023 | FLUENT


 82 NE 26th street
 Miami, FL, 33137, US

PASSED

Pages 1 of 5

PRODUCT IMAGE

SAFETY RESULTS

**Pesticides
PASSED**

**Heavy Metals
PASSED**

**Microbials
PASSED**

**Mycotoxins
PASSED**

**Residuals Solvents
NOT TESTED**

**Filtration
PASSED**

**Water Activity
PASSED**

**Moisture
PASSED**

**Terpenes
TESTED**
MISC.

Cannabinoid
PASSED

Total THC
28.151%
Total THC/Container : 281.51 mg

Total CBD
0.075%
Total CBD/Container : 0.75 mg

Total Cannabinoids
33.347%
Total Cannabinoids/Container : 333.47 mg


	D9-THC	THCA	CBD	CBDA	D8-THC	CBG	CBGA	CBN	THCV	CBDV	CBC	TOTAL CBD (DRY)	TOTAL THC (DRY)	TOTAL CANNABINOIDS (DRY)
%	0.93	31.039	ND	0.086	0.101	0.152	0.945	0.015	ND	ND	0.079	0.082	31.027	36.754
mg/unit	9.3	310.39	ND	0.86	1.01	1.52	9.45	0.15	ND	ND	0.79	0.82	310.27	367.54
LOD	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
%	%	%	%	%	%	%	%	%	%	%	%	%	%	%

 Analyzed by:
 3112, 1665, 585, 4044, 2023

 Weight:
 0.202g

 Extraction date:
 03/31/23 12:12:57

 Extracted by:
 3605,3112

 Analysis Method : SOP.T.40.031, SOP.T.30.031
 Analytical Batch : DA058131POT
 Instrument Used : DA-LC-002
 Running on : 03/31/23 13:01:09

 Reviewed On : 04/03/23 09:35:40
 Batch Date : 03/31/23 10:36:55

 Dilution : 400
 Reagent : 032923.R57; 071222.01; 033123.R01
 Consumables : 250350; CE0123; 12607-302CC-302; 61633-125C6-125E; R1KB14270
 Pipette : DA-079; DA-108; DA-078

Full Spectrum cannabinoid analysis utilizing High Performance Liquid Chromatography with UV detection in accordance with F.S. Rule 64ER20-39.



Certificate of Analysis

PASSED

FLUENT

82 NE 26th street
Miami, FL, 33137, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA30331004-007
Harvest/Lot ID: 0695 6433 5239 0312

Batch# : 0695 6433 5239 0312

Sampled : 03/30/23
Ordered : 03/30/23

Sample Size Received : 26 gram

Total Amount : 1132 units

Completed : 04/03/23 Expires: 04/03/24

Sample Method : SOP.T.20.010

Page 2 of 5

Terpenes

TESTED

Terpenes	LOD (%)	mg/unit	%	Result (%)	Terpenes	LOD (%)	mg/unit	%	Result (%)
TOTAL TERPENES	0.007	12.98	1.298		FARNESENE	0.007	0.26	0.026	
TOTAL TERPENEOL	0.007	0.38	0.038		ALPHA-HUMULENE	0.007	1.23	0.123	
ALPHA-BISABOLOL	0.007	0.53	0.053		VALENCENE	0.007	ND	ND	
ALPHA-PINENE	0.007	<0.2	<0.02		CIS-NEROLIDOL	0.007	ND	ND	
CAMPHENE	0.007	ND	ND		TRANS-NEROLIDOL	0.007	<2	<0.02	
SABINENE	0.007	ND	ND		CARYOPHYLLENE OXIDE	0.007	0.22	0.022	
BETA-PINENE	0.007	0.29	0.029		GUAIOL	0.007	ND	ND	
BETA-MYRCENE	0.007	1.03	0.103		CEDROL	0.007	ND	ND	
ALPHA-PHELLANDRENE	0.007	ND	ND		<div>Analyzed by: 2076, S&S, 4044Weight: 1.0352gExtraction date: 03/31/23 15:11:19Extracted by: 2076</div> <div>Analysis Method : SOP.T.30.061A.FL, SOP.T.40.061A.FLAnalytical Batch : DA058136TERInstrument Used : DA-GCMS-Q08Running on : 04/03/23 08:45:33Reviewed On : 04/03/23 09:35:42Batch Date : 03/31/23 11:01:53</div> <div>Dilution : 10Reagent : 121622.34Consumables : 210414634; MKCN9995; CE0123; R1KB14270Pipette : N/A</div> <div>Terpenoid testing is performed utilizing Gas Chromatography Mass Spectrometry. For all Flower samples, the Total Terpenes % is dry-weight corrected.</div>				
3-CARENE	0.007	ND	ND						
ALPHA-TERPINENE	0.007	ND	ND						
LIMONENE	0.007	1.47	0.147						
EUCALYPTOL	0.007	ND	ND						
OCIMENE	0.007	ND	ND						
GAMMA-TERPINENE	0.007	ND	ND						
SABINENE HYDRATE	0.007	ND	ND						
TERPINOLENE	0.007	ND	ND						
FENCHONE	0.007	ND	ND						
LINALOOL	0.007	0.51	0.051						
FENCHYL ALCOHOL	0.007	0.5	0.05						
ISOPULEGOL	0.007	ND	ND						
CAMPHOR	0.013	ND	ND						
ISOBORNEOL	0.007	ND	ND						
BORNEOL	0.013	ND	ND						
HEXAHYDROTHYMOL	0.007	ND	ND						
NEROL	0.007	ND	ND						
PULEGONE	0.007	ND	ND						
GERANIOL	0.007	ND	ND						
GERANYL ACETATE	0.007	ND	ND						
ALPHA-CEDRENE	0.007	ND	ND						
BETA-CARYOPHYLLENE	0.007	4.27	0.427						
Total (%)				1.298					



Certificate of Analysis

PASSED
FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

Sample : DA30331004-007

Harvest/Lot ID: 0695 6433 5239 0312

 Batch# : 0695 6433 5239
 0312

 Sampled : 03/30/23
 Ordered : 03/30/23

Sample Size Received : 26 gram

Total Amount : 1132 units

Completed : 04/03/23 Expires: 04/03/24

Sample Method : SOP.T.20.010

Page 3 of 5



Pesticides

PASSED

Pesticide	LOD	Units	Action Level	Pass/Fail	Result	Pesticide	LOD	Units	Action Level	Pass/Fail	Result
TOTAL CONTAMINANT LOAD (PESTICIDES)	0.01	ppm	5	PASS	ND	OXAMYL	0.01	ppm	0.5	PASS	ND
TOTAL DIMETHOMORPH	0.01	ppm	0.2	PASS	ND	PACLOBUTRAZOL	0.01	ppm	0.1	PASS	ND
TOTAL PERMETHRIN	0.01	ppm	0.1	PASS	ND	PHOSMET	0.01	ppm	0.1	PASS	ND
TOTAL PYRETHRINS	0.01	ppm	0.5	PASS	ND	PIPERONYL BUTOXIDE	0.01	ppm	3	PASS	ND
TOTAL SPINETORAM	0.01	ppm	0.2	PASS	ND	PRALLETHRIN	0.01	ppm	0.1	PASS	ND
TOTAL SPINOSAD	0.01	ppm	0.1	PASS	ND	PROPICONAZOLE	0.01	ppm	0.1	PASS	ND
ABAMECTIN B1A	0.01	ppm	0.1	PASS	ND	PROPOXUR	0.01	ppm	0.1	PASS	ND
ACEPHATE	0.01	ppm	0.1	PASS	ND	PYRIDABEN	0.01	ppm	0.2	PASS	ND
ACEQUINOCYL	0.01	ppm	0.1	PASS	ND	SPIROMESIFEN	0.01	ppm	0.1	PASS	ND
ACETAMIPRID	0.01	ppm	0.1	PASS	ND	SPIROTETRAMAT	0.01	ppm	0.1	PASS	ND
ALDICARB	0.01	ppm	0.1	PASS	ND	SPIROXAMINE	0.01	ppm	0.1	PASS	ND
AZOXYSTROBIN	0.01	ppm	0.1	PASS	ND	TEBUCONAZOLE	0.01	ppm	0.1	PASS	ND
BIFENAZATE	0.01	ppm	0.1	PASS	ND	THIACLOPRID	0.01	ppm	0.1	PASS	ND
BIFENTHRIN	0.01	ppm	0.1	PASS	ND	THIAMETHOXAM	0.01	ppm	0.5	PASS	ND
BOSCALID	0.01	ppm	0.1	PASS	ND	TRIFLOXYSTROBIN	0.01	ppm	0.1	PASS	ND
CARBARYL	0.01	ppm	0.5	PASS	ND	PENTACHLORONITROBENZENE (PCNB) *	0.01	PPM	0.15	PASS	ND
CARBOFURAN	0.01	ppm	0.1	PASS	ND	PARATHION-METHYL *	0.01	PPM	0.1	PASS	ND
CHLORANTRANILIPROLE	0.01	ppm	1	PASS	ND	CAPTAN *	0.07	PPM	0.7	PASS	ND
CHLORMEQUAT CHLORIDE	0.01	ppm	1	PASS	ND	CHLORDANE *	0.01	PPM	0.1	PASS	ND
CHLORPYRIFOS	0.01	ppm	0.1	PASS	ND	CHLORFENAPYR *	0.01	PPM	0.1	PASS	ND
CLOFENTEZINE	0.01	ppm	0.2	PASS	ND	CYFLUTHRIN *	0.05	PPM	0.5	PASS	ND
COUMAPHOS	0.01	ppm	0.1	PASS	ND	CYPERMETHRIN *	0.05	PPM	0.5	PASS	ND
DAMINOZIDE	0.01	ppm	0.1	PASS	ND						
DIAZINON	0.01	ppm	0.1	PASS	ND						
DICHLORVOS	0.01	ppm	0.1	PASS	ND						
DIMETHOATE	0.01	ppm	0.1	PASS	ND						
ETHOPROPHOS	0.01	ppm	0.1	PASS	ND						
ETOFENPROX	0.01	ppm	0.1	PASS	ND						
ETOXAZOLE	0.01	ppm	0.1	PASS	ND						
FENHEXAMID	0.01	ppm	0.1	PASS	ND						
FENOXYCARB	0.01	ppm	0.1	PASS	ND						
FENPYROXIMATE	0.01	ppm	0.1	PASS	ND						
FIPRONIL	0.01	ppm	0.1	PASS	ND						
FLONICAMID	0.01	ppm	0.1	PASS	ND						
FLUDIOXONIL	0.01	ppm	0.1	PASS	ND						
HEXYTHIAZOX	0.01	ppm	0.1	PASS	ND						
IMAZALIL	0.01	ppm	0.1	PASS	ND						
IMIDACLOPRID	0.01	ppm	0.4	PASS	ND						
KRESOXIM-METHYL	0.01	ppm	0.1	PASS	ND						
MALATHION	0.01	ppm	0.2	PASS	ND						
METALAXYL	0.01	ppm	0.1	PASS	ND						
METHIOCARB	0.01	ppm	0.1	PASS	ND						
METHOMYL	0.01	ppm	0.1	PASS	ND						
MEVINPHOS	0.01	ppm	0.1	PASS	ND						
MYCLOBUTANIL	0.01	ppm	0.1	PASS	ND						
NALED	0.01	ppm	0.25	PASS	ND						

 Analyzed by:
 3112, 3379, 585, 4044, 2023

 Weight:
 0.9564g

 Extraction date:
 03/31/23 15:02:02

 Extracted by:
 3379

Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.101.FL (Gainesville), SOP.T.40.102.FL (Davie)

Analytical Batch : DA058124PES

Instrument Used : DA-LCMS-003 (PES)

Running on : 03/31/23 10:44:57

Reviewed On : 04/03/23 11:54:44

Batch Date : 03/31/23 10:33:12

Dilution : 250

Reagent : 032723.R01; 032923.R26; 032923.R03; 032723.R02; 032123.R01; 032923.R01; 040521.11

Consumables : 6697075-02

Pipette : DA-093; DA-094; DA-219

Testing for agricultural agents is performed utilizing Liquid Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.

 Analyzed by:
 450, 585, 4044, 2023

 Weight:
 0.9564g

 Extraction date:
 03/31/23 15:02:02

 Extracted by:
 3379

Analysis Method : SOP.T.30.151.FL (Gainesville), SOP.T.30.151A.FL (Davie), SOP.T.40.151.FL

Analytical Batch : DA058130VOL

Instrument Used : DA-GCMS-006

Running on : 03/31/23 16:36:37

Reviewed On : 04/03/23 11:52:49

Batch Date : 03/31/23 10:36:54

Dilution : 250

Reagent : 032923.R03; 040521.11; 030923.R23; 030923.R24

Consumables : 6697075-02; 14725401

Pipette : DA-080; DA-146; DA-218

Testing for agricultural agents is performed utilizing Gas Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.



Certificate of Analysis

PASSED
FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA30331004-007

Harvest/Lot ID: 0695 6433 5239 0312

Batch# : 0695 6433 5239 0312

Sampled : 03/30/23

Ordered : 03/30/23




Sample Size Received : 26 gram

Total Amount : 1132 units

Completed : 04/03/23 **Expires:** 04/03/24

Sample Method : SOP.T.20.010

Page 4 of 5

 Microbial PASSED						 Mycotoxins PASSED					
Analyte	LOD	Units	Result	Pass / Fail	Action Level	Analyte	LOD	Units	Result	Pass / Fail	Action Level
ASPERGILLUS TERREUS			Not Present	PASS		AFLATOXIN B2	0.002	ppm	ND	PASS	0.02
ASPERGILLUS NIGER			Not Present	PASS		AFLATOXIN B1	0.002	ppm	ND	PASS	0.02
ASPERGILLUS FUMIGATUS			Not Present	PASS		OCHRATOXIN A	0.002	ppm	ND	PASS	0.02
ASPERGILLUS FLAVUS			Not Present	PASS		AFLATOXIN G1	0.002	ppm	ND	PASS	0.02
SALMONELLA SPECIFIC GENE			Not Present	PASS		AFLATOXIN G2	0.002	ppm	ND	PASS	0.02
ECOLI SHIGELLA			Not Present	PASS							
TOTAL YEAST AND MOLD	10	CFU/g	240	PASS	100000						
Analyzed by: 3336, 585, 4044, 2023 Weight: 1.9061g Extraction date: 03/31/23 11:30:41 Extracted by: 3390 Analysis Method : SOP.T.40.056C, SOP.T.40.058.FL, SOP.T.40.209.FL Analytical Batch : DA058113MIC Reviewed On : 04/03/23 09:36:43 Batch Date : 03/31/23 09:26:48 Instrument Used : PathogenDx Scanner DA-111, fisherbrand Isotemp Heat Block DA-020, fisherbrand Isotemp Heat Block DA-049, Fisher Scientific Isotemp Heat Block DA-021, APPLIED BIOSYSTEMS THERMOCYCLER DA-254 Running on : 03/31/23 12:25:33 Dilution : N/A Reagent : 011223.48; 031423.R29; 092122.07 Consumables : 7558002037 Pipette : N/A						Analyzed by: 3379, 585, 4044, 2023 Weight: 0.9564g Extraction date: 03/31/23 15:02:02 Extracted by: 3379 Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.40.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.102.FL (Davie) Analytical Batch : DA058129MYC Instrument Used : N/A Running on : 03/31/23 15:06:02 Dilution : 250 Reagent : 032723.R01; 032923.R26; 032923.R03; 032723.R02; 032123.R01; 032923.R01; 040521.11 Consumables : 6697075-02 Pipette : DA-093; DA-094; DA-219 Mycotoxins testing utilizing Liquid Chromatography with Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
 Heavy Metals PASSED											
Metal	LOD	Units	Result	Pass / Fail	Action Level						
TOTAL CONTAMINANT LOAD METALS	0.11	ppm	ND	PASS	1.1						
ARSENIC	0.02	ppm	ND	PASS	0.2						
CADMIUM	0.02	ppm	ND	PASS	0.2						
MERCURY	0.02	ppm	ND	PASS	0.2						
LEAD	0.05	ppm	ND	PASS	0.5						
Analyzed by: 1022, 585, 4044, 2023 Weight: 0.295g Extraction date: 03/31/23 10:51:01 Extracted by: 1022, 3619 Analysis Method : SOP.T.30.082.FL, SOP.T.40.082.FL Analytical Batch : DA058115HEA Instrument Used : DA-ICPMS-003 Running on : 03/31/23 14:35:25 Dilution : 50 Reagent : 031423.R28; 031423.R18; 032423.R32; 032323.R08; 032423.R30; 032423.R31; 032323.R07; 020123.02 Consumables : 179436; 210508058; 12607-302CC-302 Pipette : DA-061; DA-216 Heavy Metals analysis is performed using Inductively Coupled Plasma Mass Spectrometry in accordance with F.S. Rule 64ER20-39.											

Total yeast and mold testing is performed utilizing MPN and traditional culture based techniques in accordance with F.S. Rule 64ER20-39.



Certificate of Analysis

PASSED
FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA30331004-007

Harvest/Lot ID: 0695 6433 5239 0312

Batch# : 0695 6433 5239 0312

Sampled : 03/30/23

Ordered : 03/30/23

Sample Size Received : 26 gram

Total Amount : 1132 units

Completed : 04/03/23 **Expires:** 04/03/24

Sample Method : SOP.T.20.010

Page 5 of 5


Filth/Foreign Material
PASSED

Moisture
PASSED

Analyte	LOD	Units	Result	P/F	Action Level
Filth and Foreign Material	0.1	%	ND	PASS	1

Analyzed by: 1879, 4044	Weight: NA	Extraction date: N/A	Extracted by: N/A
-----------------------------------	----------------------	--------------------------------	-----------------------------

Analysis Method : SOP.T.40.090

Analytical Batch : DA058150FIL

Instrument Used : Filth/Foreign Material Microscope

Running on : 03/31/23 13:20:19

Reviewed On : 03/31/23 14:05:09

Batch Date : 03/31/23 12:52:44

Dilution : N/A

Reagent : N/A

Consumables : N/A

Pipette : N/A

Filth and foreign material inspection is performed by visual inspection utilizing naked eye and microscope technologies in accordance with F.S. Rule 64ER20-39.

Analyte	LOD	Units	Result	P/F	Action Level
Moisture Content	1	%	9.27	PASS	15

Analyzed by: 2926, 585, 4044	Weight: 0.506g	Extraction date: 03/31/23 15:32:03	Extracted by: 2926
--	--------------------------	--	------------------------------

Analysis Method : SOP.T.40.021

Analytical Batch : DA058100MOI

Instrument Used : DA-003 Moisture Analyzer

Running on : 03/30/23 15:55:27

Reviewed On : 03/31/23 16:03:51

Batch Date : 03/30/23 15:02:01

Dilution : N/A

Reagent : N/A

Consumables : N/A

Pipette : N/A

Moisture Content analysis utilizing loss-on-drying technology in accordance with F.S. Rule 64ER20-39.


Water Activity
PASSED

Analyte	LOD	Units	Result	P/F	Action Level
Water Activity	0.01	aw	0.515	PASS	0.65

Analyzed by: 2926, 585, 4044	Weight: 0.491g	Extraction date: 03/31/23 16:21:52	Extracted by: 2926
--	--------------------------	--	------------------------------

Analysis Method : SOP.T.40.019

Analytical Batch : DA058092WAT

Instrument Used : DA-028 Rotronic HygroPalm

Running on : 03/30/23 15:48:58

Reviewed On : 04/01/23 16:27:46

Batch Date : 03/30/23 12:13:33

Dilution : N/A

Reagent : 100522.09

Consumables : PS-14

Pipette : N/A

Water Activity is performed using a Rotronic HygroPalm HP 23-AW in accordance with F.S. Rule 64ER20-39.