



Certificate of Analysis

COMPLIANCE FOR RETAIL











Sample: DA30422003-001
Harvest/Lot ID: HYB-TT-041723-C0086
Batch#: 9242 4132 8090 8274
Cultivation Facility: Zolfo Springs Cultivation
Processing Facility: Zolfo Springs Processing
Source Facility: Zolfo Springs Cultivation
Seed to Sale# 9632 6429 1056 8643
Batch Date: 03/15/23
Sample Size Received: 31.5 gram
Total Amount: 1012 units
Retail Product Size: 3.5 gram
Ordered: 04/21/23
Sampled: 04/21/23
Completed: 04/25/23
Sampling Method: SOP.T.20.010

Apr 25, 2023 | FLUENT
82 NE 26th street
Miami, FL, 33137, US




PASSED

Pages 1 of 5

PRODUCT IMAGE	SAFETY RESULTS								MISC.
	 Pesticides PASSED	 Heavy Metals PASSED	 Microbials PASSED	 Mycotoxins PASSED	 Residuals Solvents NOT TESTED	 Filtration PASSED	 Water Activity PASSED	 Moisture PASSED	 Terpenes TESTED

	Cannabinoid	PASSED
--	--------------------	---------------



Analyzed by: 1665, 585, 4044 Weight: 0.2012g Extraction date: 04/24/23 10:50:51 Extracted by: 1665

Analysis Method : SOP.T.40.031, SOP.T.30.031
Analytical Batch : DA059191POT
Instrument Used : DA-LC-002
Analyzed Date : 04/24/23 10:55:31 Reviewed On : 04/25/23 09:59:50
Batch Date : 04/23/23 22:25:01

Dilution : 400
Reagent : 041923.R10; 071222.01; 041923.R05
Consumables : 280670723; CE0123; 61633-125C6-125E; R1KB14270
Pipette : DA-079; DA-108; DA-078

Full Spectrum cannabinoid analysis utilizing High Performance Liquid Chromatography with UV detection in accordance with F.S. Rule 64ER20-39.

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Jorge Segredo
Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164



Signature
04/25/23



Certificate of Analysis


PASSED
FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

 Sample : DA30422003-001
 Harvest/Lot ID: HYB-TT-041723-C0086

 Batch# : 9242 4132 8090
 Sample Size Received : 31.5 gram
 Total Amount : 1012 units
 Completed : 04/25/23 Expires: 04/25/24
 Sample Method : SOP.T.20.010

Page 2 of 5



Terpenes

TESTED

Terpenes	LOD (%)	mg/unit	%	Result (%)	Terpenes	LOD (%)	mg/unit	%	Result (%)
TOTAL TERPENES	0.007	118.16	3.376		FARNESENE	0.007	1.365	0.039	
TOTAL TERPINEOL	0.007	1.015	0.029		ALPHA-HUMULENE	0.007	2.03	0.058	
ALPHA-BISABOLOL	0.007	3.535	0.101		VALENCENE	0.007	ND	ND	
ALPHA-PINENE	0.007	11.9	0.34		CIS-NEROLIDOL	0.007	ND	ND	
CAMPHENE	0.007	<0.7	<0.02		TRANS-NEROLIDOL	0.007	ND	ND	
SABINENE	0.007	ND	ND		CARYOPHYLLENE OXIDE	0.007	<0.7	<0.02	
BETA-PINENE	0.007	4.095	0.117		GUAIOL	0.007	ND	ND	
BETA-MYRCENE	0.007	41.02	1.172		CEDROL	0.007	ND	ND	
ALPHA-PHELLANDRENE	0.007	ND	ND		<div>Analyzed by: 2076, 585, 4044Weight: 0.8928gExtraction date: 04/25/23 09:43:47Extracted by: 2076</div> <div>Analysis Method : SOP.T.30.061A.FL, SOP.T.40.061A.FLAnalytical Batch : DA059205TERInstrument Used : DA-GCMS-008Reviewed On : 04/25/23 16:27:48Analyzed Date : 04/25/23 09:44:32Batch Date : 04/24/23 09:39:38</div>				
3-CARENE	0.007	ND	ND						
ALPHA-TERPINENE	0.007	ND	ND						
LIMONENE	0.007	7.385	0.211						
EUCALYPTOL	0.007	ND	ND		<div>Dilution : 10Reagent : 121622.35Consumables : 210414634; MKCN9995; CE0123; R1KB14270Pipette : N/A</div> <div>Terpenoid testing is performed utilizing Gas Chromatography Mass Spectrometry. For all Flower samples, the Total Terpenes % is dry-weight corrected.</div>				
OCIMENE	0.007	4.795	0.137						
GAMMA-TERPINENE	0.007	ND	ND						
SABINENE HYDRATE	0.007	ND	ND						
TERPINOLENE	0.007	ND	ND						
FENCHONE	0.007	ND	ND						
LINALOOL	0.007	2.905	0.083						
FENCHYL ALCOHOL	0.007	1.155	0.033						
ISOPULEGOL	0.007	ND	ND						
CAMPHOR	0.013	ND	ND						
ISOBORNEOL	0.007	ND	ND						
BORNEOL	0.013	ND	ND						
HEXAHYDROTHYMOL	0.007	ND	ND						
NEROL	0.007	<0.7	<0.02						
PULEGONE	0.007	ND	ND						
GERANIOL	0.007	ND	ND						
GERANYL ACETATE	0.007	ND	ND						
ALPHA-CEDRENE	0.007	ND	ND						
BETA-CARYOPHYLLENE	0.007	7.245	0.207						
Total (%)				3.376					



Certificate of Analysis

PASSED

FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

Sample : DA30422003-001

Harvest/Lot ID: HYB-TT-041723-C0086

 Batch# : 9242 4132 8090
 8274

Sampled : 04/21/23

Ordered : 04/21/23

Sample Size Received : 31.5 gram

Total Amount : 1012 units

Completed : 04/25/23 Expires: 04/25/24

Sample Method : SOP.T.20.010

Page 3 of 5



Pesticides

PASSED

Pesticide	LOD	Units	Action Level	Pass/Fail	Result	Pesticide	LOD	Units	Action Level	Pass/Fail	Result
TOTAL CONTAMINANT LOAD (PESTICIDES)	0.01	ppm	5	PASS	ND	OXAMYL	0.01	ppm	0.5	PASS	ND
TOTAL DIMETHOMORPH	0.01	ppm	0.2	PASS	ND	PACLOBUTRAZOL	0.01	ppm	0.1	PASS	ND
TOTAL PERMETHRIN	0.01	ppm	0.1	PASS	ND	PHOSMET	0.01	ppm	0.1	PASS	ND
TOTAL PYRETHRINS	0.01	ppm	0.5	PASS	ND	PIPERONYL BUTOXIDE	0.01	ppm	3	PASS	ND
TOTAL SPINETORAM	0.01	ppm	0.2	PASS	ND	PRALLETHRIN	0.01	ppm	0.1	PASS	ND
TOTAL SPINOSAD	0.01	ppm	0.1	PASS	ND	PROPICONAZOLE	0.01	ppm	0.1	PASS	ND
ABAMECTIN B1A	0.01	ppm	0.1	PASS	ND	PROPOXUR	0.01	ppm	0.1	PASS	ND
ACEPHATE	0.01	ppm	0.1	PASS	ND	PYRIDABEN	0.01	ppm	0.2	PASS	ND
ACEQUINOCYL	0.01	ppm	0.1	PASS	ND	SPIROMESIFEN	0.01	ppm	0.1	PASS	ND
ACETAMIPRID	0.01	ppm	0.1	PASS	ND	SPIROTETRAMAT	0.01	ppm	0.1	PASS	ND
ALDICARB	0.01	ppm	0.1	PASS	ND	SPIROXAMINE	0.01	ppm	0.1	PASS	ND
AZOXYSTROBIN	0.01	ppm	0.1	PASS	ND	TEBUCONAZOLE	0.01	ppm	0.1	PASS	ND
BIFENAZATE	0.01	ppm	0.1	PASS	ND	THIACLOPRID	0.01	ppm	0.1	PASS	ND
BIFENTHRIN	0.01	ppm	0.1	PASS	ND	THIAMETHOXAM	0.01	ppm	0.5	PASS	ND
BOSCALID	0.01	ppm	0.1	PASS	ND	TRIFLOXYSTROBIN	0.01	ppm	0.1	PASS	ND
CARBARYL	0.01	ppm	0.5	PASS	ND	PENTACHLORONITROBENZENE (PCNB) *	0.01	PPM	0.15	PASS	ND
CARBOFURAN	0.01	ppm	0.1	PASS	ND	PARATHION-METHYL *	0.01	PPM	0.1	PASS	ND
CHLORANTRANILIPROLE	0.01	ppm	1	PASS	ND	CAPTAN *	0.07	PPM	0.7	PASS	ND
CHLORMEQUAT CHLORIDE	0.01	ppm	1	PASS	ND	CHLORDANE *	0.01	PPM	0.1	PASS	ND
CHLORPYRIFOS	0.01	ppm	0.1	PASS	ND	CHLORFENAPYR *	0.01	PPM	0.1	PASS	ND
CLOFENTHIZINE	0.01	ppm	0.2	PASS	ND	CYFLUTHRIN *	0.05	PPM	0.5	PASS	ND
COUMAPHOS	0.01	ppm	0.1	PASS	ND	CYPERMETHRIN *	0.05	PPM	0.5	PASS	ND
DAMINOZIDE	0.01	ppm	0.1	PASS	ND						
DIAZINON	0.01	ppm	0.1	PASS	ND	Analyzed by:	Weight:	Extraction date:	Extracted by:		
DICHLORVOS	0.01	ppm	0.1	PASS	ND	3379, 585, 4044	0.8483g	04/24/23 14:11:59	450,585		
DIMETHOATE	0.01	ppm	0.1	PASS	ND	Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.101.FL (Gainesville),					
ETHOPROPHOS	0.01	ppm	0.1	PASS	ND	SOP.T.40.102.FL (Davie)					
ETOFENPROX	0.01	ppm	0.1	PASS	ND	Analytical Batch : DA059202PES			Reviewed On : 04/25/23 09:49:05		
ETOXAZOLE	0.01	ppm	0.1	PASS	ND	Instrument Used : DA-LCMS-003 (PES)			Batch Date : 04/24/23 08:55:01		
FENHEXAMID	0.01	ppm	0.1	PASS	ND	Analyzed Date : 04/24/23 13:25:40					
FENOXYCARB	0.01	ppm	0.1	PASS	ND	Dilution : 250					
FENPYROXIMATE	0.01	ppm	0.1	PASS	ND	Reagent : 041723.R01; 042123.R02; 041823.R35; 041423.R01; 041123.R05; 041923.R01; 040521.11					
FIPRONIL	0.01	ppm	0.1	PASS	ND	Consumables : 6697075-02					
FLONICAMID	0.01	ppm	0.1	PASS	ND	Pipette : DA-093; DA-094; DA-219					
FLUDIOXONIL	0.01	ppm	0.1	PASS	ND	Testing for agricultural agents is performed utilizing Liquid Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
HEXYTHIAZOX	0.01	ppm	0.1	PASS	ND	Analyzed by:	Weight:	Extraction date:	Extracted by:		
IMAZALIL	0.01	ppm	0.1	PASS	ND	450, 585, 4044	0.8483g	04/24/23 14:11:59	450,585		
IMIDACLOPRID	0.01	ppm	0.4	PASS	ND	Analysis Method : SOP.T.30.151.FL (Gainesville), SOP.T.30.151A.FL (Davie), SOP.T.40.151.FL					
KRESOXIM-METHYL	0.01	ppm	0.1	PASS	ND	Analytical Batch : DA059204VOL			Reviewed On : 04/25/23 10:42:12		
MALATHION	0.01	ppm	0.2	PASS	ND	Instrument Used : DA-GCMS-006			Batch Date : 04/24/23 08:59:11		
METALAXYL	0.01	ppm	0.1	PASS	ND	Analyzed Date : 04/24/23 14:31:52					
METHIOCARB	0.01	ppm	0.1	PASS	ND	Dilution : 250					
METHOMYL	0.01	ppm	0.1	PASS	ND	Reagent : 041823.R35; 040521.11; 040723.R43; 040723.R44					
MEVINPHOS	0.01	ppm	0.1	PASS	ND	Consumables : 6697075-02; 14725401					
MYCLOBUTANIL	0.01	ppm	0.1	PASS	ND	Pipette : DA-080; DA-146; DA-218					
NALED	0.01	ppm	0.25	PASS	ND	Testing for agricultural agents is performed utilizing Gas Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					





Certificate of Analysis


PASSED
FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

 Sample : DA30422003-001
 Harvest/Lot ID: HYB-TT-041723-C0086
 Batch# : 9242 4132 8090
 Sample Size Received : 31.5 gram
 Total Amount : 1012 units
 Completed : 04/25/23 Expires: 04/25/24
 Sample Method : SOP.T.20.010
 8274
 Sampled : 04/21/23
 Ordered : 04/21/23

Page 4 of 5

	Microbial	PASSED		Mycotoxins	PASSED							
Analyte	LOD	Units	Result	Pass / Fail	Action Level	Analyte	LOD	Units	Result	Pass / Fail	Action Level	
ASPERGILLUS TERREUS			Not Present	PASS		AFLATOXIN B2	0.002	ppm	ND	PASS	0.02	
ASPERGILLUS NIGER			Not Present	PASS		AFLATOXIN B1	0.002	ppm	ND	PASS	0.02	
ASPERGILLUS FUMIGATUS			Not Present	PASS		OCHRATOXIN A	0.002	ppm	ND	PASS	0.02	
ASPERGILLUS FLAVUS			Not Present	PASS		AFLATOXIN G1	0.002	ppm	ND	PASS	0.02	
SALMONELLA SPECIFIC GENE			Not Present	PASS		AFLATOXIN G2	0.002	ppm	ND	PASS	0.02	
ECOLI SHIGELLA			Not Present	PASS								
TOTAL YEAST AND MOLD	10	CFU/g	540	PASS	100000	Analyzed by: 3379, 585, 4044	Weight: 0.8483g	Extraction date: 04/24/23 14:11:59		Extracted by: 450,585		
Analyzed by: 3336, 585, 4044	Weight: 1.1975g	Extraction date: 04/22/23 11:42:49		Extracted by: 3621,3336		Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.40.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.102.FL (Davie)				Reviewed On : 04/25/23 09:50:19	Batch Date : 04/24/23 08:59:07	
Analysis Method : SOP.T.40.056C, SOP.T.40.058.FL, SOP.T.40.209.FL						Analytical Batch : DA059203MYC						
Analytical Batch : DA059160MIC						Instrument Used : PathogenDx Scanner DA-111,Applied Biosystems Thermocycler DA-013,fisherbrand Isotemp Heat Block DA-020,fisherbrand Isotemp Heat Block DA-049,Fisher Scientific Isotemp Heat Block DA-021,ThermoScientific Sorvall Legend Micro 21 Centrifuge DA-018						
Instrument Used : PathogenDx Scanner DA-111,Applied Biosystems Thermocycler DA-013,fisherbrand Isotemp Heat Block DA-020,fisherbrand Isotemp Heat Block DA-049,Fisher Scientific Isotemp Heat Block DA-021,ThermoScientific Sorvall Legend Micro 21 Centrifuge DA-018						Analyzed Date : 04/22/23 16:36:29						
Dilution : N/A						Reagent : 011323.32; 041123.R27; 092122.09						
Consumables : N/A						Pipette : N/A						
Analyzed by: 3621, 3336, 585, 4044	Weight: 1.1975g	Extraction date: 04/22/23 11:42:49		Extracted by: 3621,3336								
Analysis Method : SOP.T.40.208 (Gainesville), SOP.T.40.209.FL						Reviewed On : 04/24/23 11:46:57						
Analytical Batch : DA059165TYM						Batch Date : 04/22/23 11:43:01						
Instrument Used : Incubator (25-27C) DA-096												
Analyzed Date : 04/22/23 12:06:46												
Dilution : 10												
Reagent : 011323.32												
Consumables : 007109												
Pipette : DA-212												
Total yeast and mold testing is performed utilizing MPN and traditional culture based techniques in accordance with F.S. Rule 64ER20-39.												

	Heavy Metals	PASSED			
Metal	LOD	Units	Result	Pass / Fail	Action Level
TOTAL CONTAMINANT LOAD METALS	0.08	ppm	ND	PASS	1.1
ARSENIC	0.02	ppm	ND	PASS	0.2
CADMIUM	0.02	ppm	ND	PASS	0.2
MERCURY	0.02	ppm	ND	PASS	0.2
LEAD	0.02	ppm	ND	PASS	0.5
Analyzed by: 1022, 585, 4044	Weight: 0.234g	Extraction date: 04/24/23 08:58:40		Extracted by: 3619	
Analysis Method : SOP.T.30.082.FL, SOP.T.40.082.FL					
Analytical Batch : DA059190HEA					
Instrument Used : DA-ICPMS-003					
Analyzed Date : 04/24/23 13:48:09					
Dilution : 50					
Reagent : 040623.R23; 041723.R30; 041123.R28; 040323.R21; 020123.02; 042423.R03; 042423.R01; 042423.R02					
Consumables : 179436; 210508058; 12620-307CD-307D					
Pipette : DA-061; DA-191; DA-216					
Heavy Metals analysis is performed using Inductively Coupled Plasma Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					



Certificate of Analysis

PASSED
FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

 Sample : DA30422003-001
 Harvest/Lot ID: HYB-TT-041723-C0086

 Batch# : 9242 4132 8090
 8274
 Sampled : 04/21/23
 Ordered : 04/21/23
 Sample Size Received : 31.5 gram
 Total Amount : 1012 units
 Completed : 04/25/23 Expires: 04/25/24
 Sample Method : SOP.T.20.010

Page 5 of 5


Filth/Foreign Material
PASSED

Moisture
PASSED

Analyte	LOD	Units	Result	P/F	Action Level	Analyte	LOD	Units	Result	P/F	Action Level
Filth and Foreign Material	0.1	%	ND	PASS	1	Moisture Content	1	%	13.49	PASS	15
Analyzed by: 1879, 4044 Weight: NA Extraction date: N/A Analyzed Date : 04/23/23 16:06:17						Analyzed by: 2926, 585, 4044 Weight: 0.505g Extraction date: 04/22/23 14:42:42 Analyzed Date : 04/22/23 14:40:56					
Analysis Method : SOP.T.40.090 Analytical Batch : DA059185FIL Instrument Used : Filth/Foreign Material Microscope Reviewed On : 04/23/23 16:12:10 Batch Date : 04/23/23 16:04:33						Analysis Method : SOP.T.40.021 Analytical Batch : DA059168MOI Instrument Used : DA-003 Moisture Analyzer Reviewed On : 04/22/23 17:38:56 Batch Date : 04/22/23 12:45:51					
Dilution : N/A Reagent : N/A Consumables : N/A Pipette : N/A						Dilution : N/A Reagent : 101920.06; 020123.02 Consumables : N/A Pipette : DA-066					

Filth and foreign material inspection is performed by visual inspection utilizing naked eye and microscope technologies in accordance with F.S. Rule 64ER20-39.

Moisture Content analysis utilizing loss-on-drying technology in accordance with F.S. Rule 64ER20-39.


Water Activity
PASSED

Analyte	LOD	Units	Result	P/F	Action Level
Water Activity	0.01	aw	0.637	PASS	0.65
Analyzed by: 2926, 585, 4044 Weight: 0.613g Extraction date: 04/22/23 14:28:49 Analyzed Date : 04/22/23 14:27:53					
Analysis Method : SOP.T.40.019 Analytical Batch : DA059166WAT Instrument Used : DA-028 Rotronic HygroPalm Reviewed On : 04/22/23 17:38:57 Batch Date : 04/22/23 12:38:37					
Dilution : N/A Reagent : 100522.09 Consumables : PS-14 Pipette : N/A					

Water Activity is performed using a Rotronic HygroPalm HP 23-AW in accordance with F.S. Rule 64ER20-39.