



Certificate of Analysis

COMPLIANCE FOR RETAIL











Sample: DA30617003-002
Harvest/Lot ID: BYB-BS-060823-C0092
Batch#: 0683 8307 6126 7915
Cultivation Facility: Zolfo Springs Cultivation
Processing Facility: Zolfo Springs Processing
Source Facility: Zolfo Springs Cultivation
Seed to Sale#: 2530 2464 4221 6512
Batch Date: 05/02/23
Sample Size Received: 31.5 gram
Total Amount: 1901 units
Retail Product Size: 3.5 gram
Ordered: 06/16/23
Sampled: 06/16/23
Completed: 06/20/23
Sampling Method: SOP.T.20.010

Jun 20, 2023 | FLUENT
82 NE 26th street
Miami, FL, 33137, US




PASSED

Pages 1 of 5

PRODUCT IMAGE	SAFETY RESULTS								MISC.
	 Pesticides PASSED	 Heavy Metals PASSED	 Microbials PASSED	 Mycotoxins PASSED	 Residuals Solvents NOT TESTED	 Filtration PASSED	 Water Activity PASSED	 Moisture PASSED	 Terpenes TESTED

	Cannabinoid	PASSED
--	--------------------	---------------



Analyzed by: 1665, 3112, 585, 4044 Analysis Method : SOP.T.40.031, SOP.T.30.031 Analytical Batch : DA061502POT Instrument Used : DA-LC-002 (Flower) Analyzed Date : 06/19/23 09:59:58 Dilution : 400 Reagent : 060723.R15; 071222.01; 061523.R06 Consumables : 280670723; CE0123; R1KB14270 Pipette : DA-079; DA-108; DA-078	Weight: 0.2069g	Extraction date: 06/19/23 09:55:25	Extracted by: 3112,1665
Reviewed On : 06/20/23 10:41:08 Batch Date : 06/18/23 20:09:07			

Full Spectrum cannabinoid analysis utilizing High Performance Liquid Chromatography with UV detection in accordance with F.S. Rule 64ER20-39.

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Jorge Segredo
Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164



Signature
06/20/23



Certificate of Analysis

PASSED

FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

Sample : DA30617003-002

Harvest/Lot ID: BYB-BS-060823-C0092

 Batch# : 0683 8307 6126
 7915

Sampled : 06/16/23

Ordered : 06/16/23

Sample Size Received : 31.5 gram

Total Amount : 1901 units

Completed : 06/20/23 Expires: 06/20/24

Sample Method : SOP.T.20.010

Page 2 of 5



Certificate of Analysis

PASSED

FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

Sample : DA30617003-002

Harvest/Lot ID: BYB-BS-060823-C0092

 Batch# : 0683 8307 6126
 7915

Sampled : 06/16/23

Ordered : 06/16/23

Sample Size Received : 31.5 gram

Total Amount : 1901 units

Completed : 06/20/23 Expires: 06/20/24

Sample Method : SOP.T.20.010

Page 3 of 5



Pesticides

PASSED

Pesticide	LOD	Units	Action Level	Pass/Fail	Result	Pesticide	LOD	Units	Action Level	Pass/Fail	Result
TOTAL CONTAMINANT LOAD (PESTICIDES)	0.01	ppm	5	PASS	ND	OXAMYL	0.01	ppm	0.5	PASS	ND
TOTAL DIMETHOMORPH	0.01	ppm	0.2	PASS	ND	PACLOBUTRAZOL	0.01	ppm	0.1	PASS	ND
TOTAL PERMETHRIN	0.01	ppm	0.1	PASS	ND	PHOSMET	0.01	ppm	0.1	PASS	ND
TOTAL PYRETHRINS	0.01	ppm	0.5	PASS	ND	PIPERONYL BUTOXIDE	0.01	ppm	3	PASS	ND
TOTAL SPINETORAM	0.01	ppm	0.2	PASS	ND	PRALLETHRIN	0.01	ppm	0.1	PASS	ND
TOTAL SPINOSAD	0.01	ppm	0.1	PASS	ND	PROPICONAZOLE	0.01	ppm	0.1	PASS	ND
ABAMECTIN B1A	0.01	ppm	0.1	PASS	ND	PROPOXUR	0.01	ppm	0.1	PASS	ND
ACEPHATE	0.01	ppm	0.1	PASS	ND	PYRIDABEN	0.01	ppm	0.2	PASS	ND
ACEQUINOCYL	0.01	ppm	0.1	PASS	ND	SPIROMESIFEN	0.01	ppm	0.1	PASS	ND
ACETAMIPRID	0.01	ppm	0.1	PASS	ND	SPIROTETRAMAT	0.01	ppm	0.1	PASS	ND
ALDICARB	0.01	ppm	0.1	PASS	ND	SPIROXAMINE	0.01	ppm	0.1	PASS	ND
AZOXYSTROBIN	0.01	ppm	0.1	PASS	ND	TEBUCONAZOLE	0.01	ppm	0.1	PASS	ND
BIFENAZATE	0.01	ppm	0.1	PASS	ND	THIACLOPRID	0.01	ppm	0.1	PASS	ND
BIFENTHRIN	0.01	ppm	0.1	PASS	ND	THIAMETHOXAM	0.01	ppm	0.5	PASS	ND
BOSCALID	0.01	ppm	0.1	PASS	ND	TRIFLOXYSTROBIN	0.01	ppm	0.1	PASS	ND
CARBARYL	0.01	ppm	0.5	PASS	ND	PENTACHLORONITROBENZENE (PCNB) *	0.01	PPM	0.15	PASS	ND
CARBOFURAN	0.01	ppm	0.1	PASS	ND	PARATHION-METHYL *	0.01	PPM	0.1	PASS	ND
CHLORANTRANILIPROLE	0.01	ppm	1	PASS	ND	CAPTAN *	0.07	PPM	0.7	PASS	ND
CHLORMEQUAT CHLORIDE	0.01	ppm	1	PASS	ND	CHLORDANE *	0.01	PPM	0.1	PASS	ND
CHLORPYRIFOS	0.01	ppm	0.1	PASS	ND	CHLORFENAPYR *	0.01	PPM	0.1	PASS	ND
CLOFENTHIZINE	0.01	ppm	0.2	PASS	ND	CYFLUTHRIN *	0.05	PPM	0.5	PASS	ND
COUMAPHOS	0.01	ppm	0.1	PASS	ND	CYPERMETHRIN *	0.05	PPM	0.5	PASS	ND
DAMINOZIDE	0.01	ppm	0.1	PASS	ND						
DIAZINON	0.01	ppm	0.1	PASS	ND	Analyzed by:	Weight:	Extraction date:	Extracted by:		
DICHLORVOS	0.01	ppm	0.1	PASS	ND	3379, 585, 4044	1.0207g	06/18/23 17:58:42	4056		
DIMETHOATE	0.01	ppm	0.1	PASS	ND	Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.101.FL (Gainesville),					
ETHOPROPHOS	0.01	ppm	0.1	PASS	ND	SOP.T.40.102.FL (Davie)					
ETOFENPROX	0.01	ppm	0.1	PASS	ND	Analytical Batch : DA061498PES			Reviewed On : 06/20/23 11:43:41		
ETOXAZOLE	0.01	ppm	0.1	PASS	ND	Instrument Used : DA-LCMS-003 (PES)			Batch Date : 06/18/23 13:04:04		
FENHEXAMID	0.01	ppm	0.1	PASS	ND	Analyzed Date : 06/19/23 12:59:40					
FENOXYCARB	0.01	ppm	0.1	PASS	ND	Dilution : 250					
FENPYROXIMATE	0.01	ppm	0.1	PASS	ND	Reagent : 061423.R23; 040521.11; 061223.R02; 061223.R05; 061223.R04; 060523.R26; 061423.R08					
FIPRONIL	0.01	ppm	0.1	PASS	ND	Consumables : 6697075-02					
FLONICAMID	0.01	ppm	0.1	PASS	ND	Pipette : DA-093; DA-094; DA-219					
FLUDIOXONIL	0.01	ppm	0.1	PASS	ND	Testing for agricultural agents is performed utilizing Liquid Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
HEXYTHIAZOX	0.01	ppm	0.1	PASS	ND	Analyzed by:	Weight:	Extraction date:	Extracted by:		
IMAZALIL	0.01	ppm	0.1	PASS	ND	450, 795, 585, 4044	1.0207g	06/18/23 17:58:42	4056		
IMIDACLOPRID	0.01	ppm	0.4	PASS	ND	Analysis Method : SOP.T.30.151.FL (Gainesville), SOP.T.30.151A.FL (Davie), SOP.T.40.151.FL					
KRESOXIM-METHYL	0.01	ppm	0.1	PASS	ND	Analytical Batch : DA061499VOL			Reviewed On : 06/20/23 16:39:11		
MALATHION	0.01	ppm	0.2	PASS	ND	Instrument Used : DA-GCMS-006			Batch Date : 06/18/23 13:05:01		
METALAXYL	0.01	ppm	0.1	PASS	ND	Analyzed Date : 06/19/23 13:12:48					
METHIOCARB	0.01	ppm	0.1	PASS	ND	Dilution : 250					
METHOMYL	0.01	ppm	0.1	PASS	ND	Reagent : 061423.R23; 040521.11; 061223.R25; 061223.R24					
MEVINPHOS	0.01	ppm	0.1	PASS	ND	Consumables : 6697075-02; 14725401					
MYCLOBUTANIL	0.01	ppm	0.1	PASS	ND	Pipette : DA-080; DA-146; DA-218					
NALED	0.01	ppm	0.25	PASS	ND	Testing for agricultural agents is performed utilizing Gas Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					



Certificate of Analysis




PASSED
FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

 Sample : DA30617003-002
 Harvest/Lot ID: BYB-BS-060823-C0092

 Batch# : 0683 8307 6126 Sample Size Received : 31.5 gram
 7915 Total Amount : 1901 units
 Sampled : 06/16/23 Completed : 06/20/23 Expires: 06/20/24
 Ordered : 06/16/23 Sample Method : SOP.T.20.010

Page 4 of 5

<div></div> <div>Microbial</div>						<div></div> <div>Mycotoxins</div>						<div>PASSED</div>					
Analyte	LOD	Units	Result	Pass / Fail	Action Level	Analyte	LOD	Units	Result	Pass / Fail	Action Level						
ASPERGILLUS TERREUS			Not Present	PASS		AFLATOXIN B2	0.002	ppm	ND	PASS	0.02						
ASPERGILLUS NIGER			Not Present	PASS		AFLATOXIN B1	0.002	ppm	ND	PASS	0.02						
ASPERGILLUS FUMIGATUS			Not Present	PASS		OCHRATOXIN A	0.002	ppm	ND	PASS	0.02						
ASPERGILLUS FLAVUS			Not Present	PASS		AFLATOXIN G1	0.002	ppm	ND	PASS	0.02						
SALMONELLA SPECIFIC GENE			Not Present	PASS		AFLATOXIN G2	0.002	ppm	ND	PASS	0.02						
ECOLI SHIGELLA			Not Present	PASS													
TOTAL YEAST AND MOLD	10	CFU/g	960	PASS	100000	Analyzed by: 3379, 585, 4044	Weight: 1.0207g	Extraction date: 06/18/23 17:58:42		Extracted by: 4056							
Analyzed by: 3621, 585, 4044	Weight: 1.1915g	Extraction date: 06/17/23 15:53:51		Extracted by: 3621,3390		Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.40.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.102.FL (Davie)											
Analysis Method : SOP.T.40.056C, SOP.T.40.058.FL, SOP.T.40.209.FL						Analytical Batch : DA061500MYC					Reviewed On : 06/20/23 10:26:40						
Analytical Batch : DA061468MIC					Reviewed On : 06/20/23 16:33:11	Instrument Used : N/A					Batch Date : 06/18/23 13:07:05						
					Batch Date : 06/17/23	Analyzed Date : 06/19/23 12:59:18											
Instrument Used : PathogenDx Scanner DA-111,Applied Biosystems Thermocycler DA-010,fisherbrand Isotemp Heat Block DA-020,fisherbrand Isotemp Heat Block DA-049,Fisher Scientific Isotemp Heat Block DA-021						Dilution : 250											
Analyzed Date : 06/17/23 17:08:34						Reagent : 061423.R23; 040521.11; 061223.R02; 061623.R05; 061223.R04; 060523.R26; 061423.R08											
						Consumables : 6697075-02											
						Pipette : DA-093; DA-094; DA-219											
Dilution : N/A						Mycotoxins testing utilizing Liquid Chromatography with Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.											
Reagent : 031523.18; 052323.R22; 020823.16; 092122.09																	
Consumables : 7562003041																	
Pipette : N/A																	
Analyzed by: 3621, 3702, 585, 4044	Weight: 1.1915g	Extraction date: N/A		Extracted by: 3621,3390		<div></div> <div>Heavy Metals</div>		<div>PASSED</div>									
Analysis Method : SOP.T.40.208 (Gainesville), SOP.T.40.209.FL						Metal	LOD	Units	Result	Pass / Fail	Action Level						
Analytical Batch : DA061469TYM						TOTAL CONTAMINANT LOAD METALS	0.08	ppm	ND	PASS	1.1						
Instrument Used : Incubator (25-27C) DA-096						ARSENIC	0.02	ppm	ND	PASS	0.2						
Analyzed Date : 06/17/23 17:12:29						CADMIUM	0.02	ppm	ND	PASS	0.2						
						MERCURY	0.02	ppm	ND	PASS	0.2						
						LEAD	0.02	ppm	ND	PASS	0.5						
Total yeast and mold testing is performed utilizing MPN and traditional culture based techniques in accordance with F.S. Rule 64ER20-39.						Analyzed by: 1022, 585, 4044	Weight: 0.21g	Extraction date: 06/17/23 13:12:35		Extracted by: 3807,3619							
						Analysis Method : SOP.T.30.082.FL, SOP.T.40.082.FL											
						Analytical Batch : DA061486HEA						Reviewed On : 06/20/23 10:23:09					
						Instrument Used : DA-ICPMS-003						Batch Date : 06/17/23 12:39:20					
						Analyzed Date : 06/19/23 12:56:00											
						Dilution : 50											
						Reagent : 061523.R17; 042623.R82; 061623.R25; 061623.R06; 061623.R23; 061623.R24; 052523.R15; 061423.R46											
						Consumables : 179436; 15021042; 210508058											
						Pipette : DA-061; DA-191; DA-216											
						Heavy Metals analysis is performed using Inductively Coupled Plasma Mass Spectrometry in accordance with F.S. Rule 64ER20-39.											



Certificate of Analysis

PASSED
FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

 Sample : DA30617003-002
 Harvest/Lot ID: BYB-BS-060823-C0092

 Batch# : 0683 8307 6126
 Sample Size Received : 31.5 gram
 Total Amount : 1901 units
 Completed : 06/20/23 Expires: 06/20/24
 Sampled : 06/16/23
 Ordered : 06/16/23
 Sample Method : SOP.T.20.010

Page 5 of 5


Filth/Foreign Material
PASSED

Moisture
PASSED

Analyte	LOD	Units	Result	P/F	Action Level	Analyte	LOD	Units	Result	P/F	Action Level
Filth and Foreign Material	0.1	%	ND	PASS	1	Moisture Content	1	%	14.53	PASS	15
Analyzed by: 1879, 4044 Weight: NA Extraction date: N/A Analyzed Date: 06/17/23 18:19:19						Analyzed by: 1879, 4056, 585, 4044 Weight: 1.308g Extraction date: 06/17/23 14:46:27 Analyzed Date: 06/17/23 13:59:35					
Analysis Method : SOP.T.40.090 Analytical Batch : DA061489FIL Instrument Used : Filth/Foreign Material Microscope Reviewed On : 06/17/23 18:31:32 Batch Date : 06/17/23 18:15:28						Analysis Method : SOP.T.40.021 Analytical Batch : DA061476MOI Instrument Used : DA-003 Moisture Analyzer Reviewed On : 06/17/23 15:17:11 Batch Date : 06/17/23 11:46:41					
Dilution : N/A Reagent : N/A Consumables : N/A Pipette : N/A						Dilution : N/A Reagent : 101920.06; 020123.02 Consumables : N/A Pipette : DA-066					

Filth and foreign material inspection is performed by visual inspection utilizing naked eye and microscope technologies in accordance with F.S. Rule 64ER20-39.

Moisture Content analysis utilizing loss-on-drying technology in accordance with F.S. Rule 64ER20-39.


Water Activity
PASSED

Analyte	LOD	Units	Result	P/F	Action Level
Water Activity	0.01	aw	0.58	PASS	0.65
Analyzed by: 4056, 585, 4044 Weight: 1.362g Extraction date: 06/17/23 15:36:27 Analyzed Date: N/A					
Analysis Method : SOP.T.40.019 Analytical Batch : DA061473WAT Instrument Used : DA-028 Rotronic HygroPalm Reviewed On : 06/19/23 11:26:12 Batch Date : 06/17/23 11:35:14					
Dilution : N/A Reagent : 050923.03 Consumables : PS-14 Pipette : N/A					

Water Activity is performed using a Rotronic HygroPalm HP 23-AW in accordance with F.S. Rule 64ER20-39.