



Certificate of Analysis

PASSED

FLUENT

82 NE 26th street
Miami, FL, 33137, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA30713005-002
Harvest/Lot ID: HYB-SFVXB-070323-C0097

Batch# : 5600 6416 1303 Sample Size Received : 31.5 gram
2251 Total Amount : 1861 units
Sampled : 07/12/23 Completed : 07/15/23 Expires: 07/15/24
Ordered : 07/12/23 Sample Method : SOP.T.20.010

Page 2 of 5

Terpenes

TESTED

Terpenes	LOD (%)	mg/unit	%	Result (%)	Terpenes	LOD (%)	mg/unit	%	Result (%)
TOTAL TERPENES	0.02	82.32	2.352		FARNESENE		0.245	0.007	
TOTAL TERPINEOL	0.02	1.54	0.044		ALPHA-HUMULENE	0.02	3.955	0.113	
ALPHA-BISABOLOL	0.02	2.625	0.075		VALENCENE	0.02	<0.7	<0.02	
ALPHA-PINENE	0.02	2.135	0.061		CIS-NEROLIDOL	0.02	ND	ND	
CAMPHENE	0.02	0.77	0.022		TRANS-NEROLIDOL	0.02	ND	ND	
SABINENE	0.02	ND	ND		CARYOPHYLLENE OXIDE	0.02	0.875	0.025	
BETA-PINENE	0.02	3.08	0.088		GUAJOL	0.02	2.695	0.077	
BETA-MYRCENE	0.02	20.405	0.583		CEDROL	0.02	ND	ND	
ALPHA-PHELLANDRENE	0.02	ND	ND		Analyzed by: 2076, 585, 1440	Weight: 0.9228g	Extraction date: 07/14/23 09:18:26	Extracted by: 2076	
3-CARENE	0.02	ND	ND		Analysis Method : SOP.T.30.061A.FL, SOP.T.40.061A.FL				
ALPHA-TERPINENE	0.02	ND	ND		Analytical Batch : DA062275TER			Reviewed On : 07/15/23 14:55:49	
LIMONENE	0.02	13.09	0.374		Instrument Used : DA-GCMS-004			Batch Date : 07/13/23 09:41:01	
EUCALYPTOL	0.02	<0.7	<0.02		Analyzed Date : N/A				
OCIMENE	0.02	<0.7	<0.02		Dilution : 10				
GAMMA-TERPINENE	0.02	ND	ND		Reagent : 121622.26				
SABINENE HYDRATE	0.02	ND	ND		Consumables : 210414634; MKCN9995; CE0123; R1KB14270				
TERPINOLENE	0.02	ND	ND		Pipette : N/A				
FENCHONE	0.04	<1.4	<0.04		Terpenoid testing is performed utilizing Gas Chromatography Mass Spectrometry. For all Flower samples, the Total Terpenes % is dry-weight corrected.				
LINALOOL	0.02	5.95	0.17						
FENCHYL ALCOHOL	0.02	1.855	0.053						
ISOPULEGOL	0.02	<0.7	<0.02						
CAMPHOR	0.06	<2.1	<0.06						
ISOBORNEOL	0.02	ND	ND						
BORNEOL	0.04	<1.4	<0.04						
HEXAHYDROTHYMOL	0.02	ND	ND						
NEROL	0.02	ND	ND						
PULEGONE	0.02	ND	ND						
GERANIOL	0.02	<0.7	<0.02						
GERANYL ACETATE	0.02	ND	ND						
ALPHA-CEDRENE	0.02	ND	ND						
BETA-CARYOPHYLLENE	0.02	13.51	0.386						
Total (%)			2.352						



Certificate of Analysis

PASSED


FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

 Sample : DA30713005-002
 Harvest/Lot ID: HYB-SFVXB-070323-C0097

 Batch# : 5600 6416 1303 Sample Size Received : 31.5 gram
 2251 Total Amount : 1861 units
 Sampled : 07/12/23 Completed : 07/15/23 Expires: 07/15/24
 Ordered : 07/12/23 Sample Method : SOP.T.20.010

Page 3 of 5

<div><div></div><div>Pesticides</div></div>						PASSED					
Pesticide	LOD	Units	Action Level	Pass/Fail	Result	Pesticide	LOD	Units	Action Level	Pass/Fail	Result
TOTAL CONTAMINANT LOAD (PESTICIDES)	0.01	ppm	5	PASS	ND	OXAMYL	0.01	ppm	0.5	PASS	ND
TOTAL DIMETHOMORPH	0.01	ppm	0.2	PASS	ND	PACLOBUTRAZOL	0.01	ppm	0.1	PASS	ND
TOTAL PERMETHRIN	0.01	ppm	0.1	PASS	ND	PHOSMET	0.01	ppm	0.1	PASS	ND
TOTAL PYRETHRINS	0.01	ppm	0.5	PASS	ND	PIPERONYL BUTOXIDE	0.01	ppm	3	PASS	ND
TOTAL SPINETORAM	0.01	ppm	0.2	PASS	ND	PRALLETHRIN	0.01	ppm	0.1	PASS	ND
TOTAL SPINOSAD	0.01	ppm	0.1	PASS	ND	PROPICONAZOLE	0.01	ppm	0.1	PASS	ND
ABAMECTIN B1A	0.01	ppm	0.1	PASS	ND	PROPOXUR	0.01	ppm	0.1	PASS	ND
ACEPHATE	0.01	ppm	0.1	PASS	ND	PYRIDABEN	0.01	ppm	0.2	PASS	ND
ACEQUINOCYL	0.01	ppm	0.1	PASS	ND	SPIROMESIFEN	0.01	ppm	0.1	PASS	ND
ACETAMIPRID	0.01	ppm	0.1	PASS	ND	SPIROTETRAMAT	0.01	ppm	0.1	PASS	ND
ALDICARB	0.01	ppm	0.1	PASS	ND	SPIROXAMINE	0.01	ppm	0.1	PASS	ND
AZOXYSTROBIN	0.01	ppm	0.1	PASS	ND	TEBUCONAZOLE	0.01	ppm	0.1	PASS	ND
BIFENAZATE	0.01	ppm	0.1	PASS	ND	THIACLOPRID	0.01	ppm	0.1	PASS	ND
BIFENTHRIN	0.01	ppm	0.1	PASS	ND	THIAMETHOXAM	0.01	ppm	0.5	PASS	ND
BOSCALID	0.01	ppm	0.1	PASS	ND	TRIFLOXYSTROBIN	0.01	ppm	0.1	PASS	ND
CARBARYL	0.01	ppm	0.5	PASS	ND	PENTACHLORONITROBENZENE (PCNB) *	0.05	PPM	0.15	PASS	ND
CARBOFURAN	0.01	ppm	0.1	PASS	ND	PARATHION-METHYL *	0.05	PPM	0.1	PASS	ND
CHLORANTRANILIPROLE	0.01	ppm	1	PASS	ND	CAPTAN *	0.35	PPM	0.7	PASS	ND
CHLORMEQUAT CHLORIDE	0.01	ppm	1	PASS	ND	CHLORDANE *	0.05	PPM	0.1	PASS	ND
CHLORPYRIFOS	0.01	ppm	0.1	PASS	ND	CHLORFENAPYR *	0.05	PPM	0.1	PASS	ND
CLOFENTEZINE	0.01	ppm	0.2	PASS	ND	CYFLUTHRIN *	0.25	PPM	0.5	PASS	ND
COUMAPHOS	0.01	ppm	0.1	PASS	ND	CYPERMETHRIN *	0.25	PPM	0.5	PASS	ND
DAMINOZIDE	0.01	ppm	0.1	PASS	ND	Analyzed by:	Weight:	Extraction date:	Extracted by:		
DIAZINON	0.01	ppm	0.1	PASS	ND	3379, 585, 1440	0.9584g	07/13/23 15:25:36	450,585		
DICHLORVOS	0.01	ppm	0.1	PASS	ND	Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.101.FL (Gainesville),					
DIMETHOATE	0.01	ppm	0.1	PASS	ND	SOP.T.40.102.FL (Davie)					
ETHOPROPHOS	0.01	ppm	0.1	PASS	ND	Analytical Batch : DA062291PES				Reviewed On : 07/15/23 19:41:24	
ETOFENPROX	0.01	ppm	0.1	PASS	ND	Instrument Used : DA-LCMS-003 (PES)				Batch Date : 07/13/23 10:40:17	
ETOXAZOLE	0.01	ppm	0.1	PASS	ND	Analyzed Date : N/A					
FENHEXAMID	0.01	ppm	0.1	PASS	ND	Dilution : 250					
FENOXYCARB	0.01	ppm	0.1	PASS	ND	Reagent : 071023.R04; 071123.R18; 071323.R03; 070723.R01; 060523.R26; 071323.R01; 040521.11					
FENPYROXIMATE	0.01	ppm	0.1	PASS	ND	Consumables : 326250IW					
FIPRONIL	0.01	ppm	0.1	PASS	ND	Pipette : DA-093; DA-094; DA-219					
FLONICAMID	0.01	ppm	0.1	PASS	ND	Testing for agricultural agents is performed utilizing Liquid Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
FLUDIOXONIL	0.01	ppm	0.1	PASS	ND	Analyzed by:	Weight:	Extraction date:	Extracted by:		
HEXYTHIAZOX	0.01	ppm	0.1	PASS	ND	450, 585, 1440	0.9584g	07/13/23 15:25:36	450,585		
IMAZALIL	0.01	ppm	0.1	PASS	ND	Analysis Method : SOP.T.30.151.FL (Gainesville), SOP.T.30.151A.FL (Davie), SOP.T.40.151.FL					
IMIDACLOPRID	0.01	ppm	0.4	PASS	ND	Analytical Batch : DA062293VOL				Reviewed On : 07/15/23 14:51:51	
KRESOXIM-METHYL	0.01	ppm	0.1	PASS	ND	Instrument Used : DA-GCMS-001				Batch Date : 07/13/23 10:41:45	
MALATHION	0.01	ppm	0.2	PASS	ND	Analyzed Date : 07/13/23 16:27:19					
METALAXYL	0.01	ppm	0.1	PASS	ND	Dilution : 250					
METHIOCARB	0.01	ppm	0.1	PASS	ND	Reagent : 071323.R03; 040521.11; 071123.R21; 071123.R22					
METHOMYL	0.01	ppm	0.1	PASS	ND	Consumables : 326250IW; 14725401					
MEVINPHOS	0.01	ppm	0.1	PASS	ND	Pipette : DA-080; DA-146; DA-218					
MYCLOBUTANIL	0.01	ppm	0.1	PASS	ND	Testing for agricultural agents is performed utilizing Gas Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
NALED	0.01	ppm	0.25	PASS	ND						






Certificate of Analysis

PASSED
FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

 Sample : DA30713005-002
 Harvest/Lot ID: HYB-SFVXB-070323-C0097
 Batch# : 5600 6416 1303
 Sample Size Received : 31.5 gram
 Total Amount : 1861 units
 Completed : 07/15/23 Expires: 07/15/24
 Sample Method : SOP.T.20.010
 Sampled : 07/12/23
 Ordered : 07/12/23

Page 4 of 5

<div></div> <div>Microbial</div> <div>PASSED</div>						<div></div> <div>Mycotoxins</div> <div>PASSED</div>					
Analyte	LOD	Units	Result	Pass / Fail	Action Level	Analyte	LOD	Units	Result	Pass / Fail	Action Level
ASPERGILLUS TERREUS			Not Present	PASS		AFLATOXIN B2	0.002	ppm	ND	PASS	0.02
ASPERGILLUS NIGER			Not Present	PASS		AFLATOXIN B1	0.002	ppm	ND	PASS	0.02
ASPERGILLUS FUMIGATUS			Not Present	PASS		OCHRATOXIN A	0.002	ppm	ND	PASS	0.02
ASPERGILLUS FLAVUS			Not Present	PASS		AFLATOXIN G1	0.002	ppm	ND	PASS	0.02
SALMONELLA SPECIFIC GENE			Not Present	PASS		AFLATOXIN G2	0.002	ppm	ND	PASS	0.02
ECOLI SHIGELLA			Not Present	PASS							
TOTAL YEAST AND MOLD	10	CFU/g	40	PASS	100000						
Analyzed by: 3390, 3336, 585, 1440 Weight: 0.8417g Extraction date: 07/13/23 10:00:36 Extracted by: 3621,3390						Analyzed by: 3379, 585, 1440 Weight: 0.9584g Extraction date: 07/13/23 15:25:36 Extracted by: 450,585					
Analysis Method : SOP.T.40.056C, SOP.T.40.058.FL, SOP.T.40.209.FL Analytical Batch : DA062269MIC Instrument Used : PathogenDx Scanner DA-111,Applied Biosystems MiniAmp Thermocycler DA-190,fisherbrand Isotemp Heat Block DA-020,fisherbrand Isotemp Heat Block DA-049,Fisher Scientific Isotemp Heat Block DA-021 Analyzed Date : 07/13/23 15:36:23						Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.40.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.102.FL (Davie) Analytical Batch : DA062292MYC Instrument Used : N/A Analyzed Date : N/A Dilution : 250 Reagent : 071023.R04; 071123.R18; 071323.R03; 070723.R01; 060523.R26; 071323.R01; 040521.11 Consumables : 326250IW Pipette : DA-093; DA-094; DA-219					
Dilution : N/A Reagent : 050223.50; 062323.R18; 020823.14; 092122.09 Consumables : 7562003045 Pipette : N/A						Mycotoxins testing utilizing Liquid Chromatography with Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
Analyzed by: 3336, 585, 1440 Weight: 0.8417g Extraction date: N/A Extracted by: 3621,3336						<div></div> <div>Heavy Metals</div> <div>PASSED</div>					
Analysis Method : SOP.T.40.208 (Gainesville), SOP.T.40.209.FL Analytical Batch : DA062276TYM Instrument Used : Incubator (25-27C) DA-096 Analyzed Date : 07/13/23 16:08:29						Metal LOD Units Result Pass / Fail Action Level					
Dilution : 10 Reagent : 050223.50; 070523.R46 Consumables : N/A Pipette : N/A						TOTAL CONTAMINANT LOAD METALS 0.08 ppm ND PASS 1.1					
Total yeast and mold testing is performed utilizing MPN and traditional culture based techniques in accordance with F.S. Rule 64ER20-39.						ARSENIC 0.02 ppm ND PASS 0.2					
						CADMIUM 0.02 ppm ND PASS 0.2					
						MERCURY 0.02 ppm ND PASS 0.2					
						LEAD 0.02 ppm ND PASS 0.5					
Analyzed by: 1022, 585, 1440 Weight: 0.2051g Extraction date: 07/13/23 11:03:25 Extracted by: 3619						Analyzed by: 1022, 585, 1440 Weight: 0.2051g Extraction date: 07/13/23 11:03:25 Extracted by: 3619					
Analysis Method : SOP.T.30.082.FL, SOP.T.40.082.FL Analytical Batch : DA062284HEA Instrument Used : DA-ICPMS-003 Analyzed Date : 07/13/23 17:13:34						Analysis Method : SOP.T.30.082.FL, SOP.T.40.082.FL Analytical Batch : DA062284HEA Instrument Used : DA-ICPMS-003 Analyzed Date : 07/13/23 17:13:34					
Dilution : 50 Reagent : 061523.R17; 062723.R18; 070723.R17; 071123.R17; 070723.R15; 070723.R16; 070723.R18; 071023.01; 062823.R15 Consumables : 179436; 15021042; 210508058 Pipette : DA-061; DA-191; DA-216						Dilution : 50 Reagent : 061523.R17; 062723.R18; 070723.R17; 071123.R17; 070723.R15; 070723.R16; 070723.R18; 071023.01; 062823.R15 Consumables : 179436; 15021042; 210508058 Pipette : DA-061; DA-191; DA-216					
Heavy Metals analysis is performed using Inductively Coupled Plasma Mass Spectrometry in accordance with F.S. Rule 64ER20-39.						Heavy Metals analysis is performed using Inductively Coupled Plasma Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					



Certificate of Analysis

PASSED
FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

 Sample : DA30713005-002
 Harvest/Lot ID: HYB-SFVXB-070323-C0097

 Batch# : 5600 6416 1303 Sample Size Received : 31.5 gram
 2251 Total Amount : 1861 units
 Sampled : 07/12/23 Completed : 07/15/23 Expires: 07/15/24
 Ordered : 07/12/23 Sample Method : SOP.T.20.010

Page 5 of 5


Filth/Foreign Material
PASSED

Moisture
PASSED

Analyte	LOD	Units	Result	P/F	Action Level	Analyte	LOD	Units	Result	P/F	Action Level
Filth and Foreign Material	0.1	%	ND	PASS	1	Moisture Content	1	%	11.66	PASS	15
Analyzed by: 585, 1440 Weight: NA Extraction date: N/A Extracted by: N/A						Analyzed by: 3807, 585, 1440 Weight: 0.501g Extraction date: 07/13/23 14:47:54 Extracted by: 3807					
Analysis Method : SOP.T.40.090 Analytical Batch : DA062384FIL Instrument Used : Filth/Foreign Material Microscope Analyzed Date : N/A			Reviewed On : 07/15/23 14:42:51 Batch Date : 07/15/23 14:31:52			Analysis Method : SOP.T.40.021 Analytical Batch : DA062297MOI Instrument Used : DA-003 Moisture Analyzer Analyzed Date : N/A			Reviewed On : 07/14/23 14:01:47 Batch Date : 07/13/23 10:47:29		
Dilution : N/A Reagent : N/A Consumables : N/A Pipette : N/A						Dilution : N/A Reagent : 101920.06; 020123.02 Consumables : N/A Pipette : DA-066					

Filth and foreign material inspection is performed by visual inspection utilizing naked eye and microscope technologies in accordance with F.S. Rule 64ER20-39.

Moisture Content analysis utilizing loss-on-drying technology in accordance with F.S. Rule 64ER20-39.


Water Activity
PASSED

Analyte	LOD	Units	Result	P/F	Action Level
Water Activity	0.1	aw	0.557	PASS	0.65
Analyzed by: 3807, 585, 1440 Weight: 0.6g Extraction date: 07/13/23 15:00:41 Extracted by: 3807					
Analysis Method : SOP.T.40.019 Analytical Batch : DA062299WAT Instrument Used : DA-028 Rotronic HygroPalm Analyzed Date : N/A			Reviewed On : 07/14/23 14:01:48 Batch Date : 07/13/23 10:49:30		
Dilution : N/A Reagent : 050923.04 Consumables : PS-14 Pipette : N/A					

Water Activity is performed using a Rotronic HygroPalm HP 23-AW in accordance with F.S. Rule 64ER20-39.