



Certificate of Analysis

COMPLIANCE FOR RETAIL

Sample: DA31130006-001
 Harvest/Lot ID: 8923 3224 6475 6903
 Batch#: 8923 3224 6475 6903
 Cultivation Facility: Zolfo Springs Cultivation
 Processing Facility: Zolfo Springs Processing
 Source Facility: Zolfo Springs Cultivation
 Seed to Sale# 0934 6458 5765 5919
 Batch Date: 10/20/23
 Sample Size Received: 31.5 gram
 Total Amount: 544 units
 Retail Product Size: 3.5 gram
 Ordered: 11/29/23
 Sampled: 11/30/23
 Completed: 12/02/23
 Sampling Method: SOP.T.20.010

Dec 02, 2023 | FLUENT


 82 NE 26th street
 Miami, FL, 33137, US

PASSED

Pages 1 of 5

PRODUCT IMAGE

SAFETY RESULTS

 Pesticides
PASSED

 Heavy Metals
PASSED

 Microbials
PASSED

 Mycotoxins
PASSED

 Residuals Solvents
 NOT TESTED

 Filtration
PASSED

 Water Activity
PASSED

 Moisture
PASSED

 Terpenes
TESTED
MISC.

Cannabinoid
PASSED

Total THC
26.975%
 Dry Weight

Total CBD
0.058%
 Dry Weight

Total Cannabinoids
31.578%
 Dry Weight

	D9-THC	THCA	CBD	CBDA	D8-THC	CBG	CBGA	CBN	THCV	CBDV	CBC
%	0.711	25.514	ND	0.058	0.044	0.068	0.585	ND	ND	ND	0.045
mg/unit	24.885	892.99	ND	2.03	1.54	2.38	20.475	ND	ND	ND	1.575
LOD	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
	%	%	%	%	%	%	%	%	%	%	%

Total THC
23.086%
 808.01 mg /Container

Total CBD
0.05%
 1.75 mg /Container

Total Cannabinoids
27.025%
 945.875 mg /Container

As Received

 Analyzed by:
 3335, 585, 3963

 Weight:
 0.2026g

 Extraction date:
 11/30/23 15:20:32

 Extracted by:
 3335

Analysis Method : SOP.T.40.031, SOP.T.30.031

Analytical Batch : DA066877POT

Instrument Used : DA-LC-002

Analyzed Date : 11/30/23 16:00:26

Reviewed On : 12/01/23 09:56:26

Batch Date : 11/30/23 11:17:21

Dilution : 400

Reagent : 112923.R03; 060723.24; 110723.R05

Consumables : 947.109; CE0123; 12594-247CD-247C; R1KB14270

Pipette : DA-079; DA-108; DA-078

Full Spectrum cannabinoid analysis utilizing High Performance Liquid Chromatography with UV detection in accordance with F.S. Rule 64ER20-39.

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino

Lab Director

 State License # CMTL-0002
 ISO 17025 Accreditation # ISO/IEC
 17025:2017 Accreditation PJLA-
 Testing 97164

 Signature
 12/02/23



4131 SW 47th AVENUE SUITE 1408
DAVIE, FL, 33314, US
(954) 368-7664

Kaycha Labs

FTH-Bubbly Haze #340 WF 3.5g(1/8oz)
FTH-Bubbly Haze #340
Matrix : Flower
Type: Flower-Cured



Certificate of Analysis

PASSED

FLUENT

82 NE 26th street
Miami, FL, 33137, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA31130006-001

Harvest/Lot ID: 8923 3224 6475 6903

Batch# : 8923 3224 6475
6903

Sampled : 11/30/23
Ordered : 11/30/23

Sample Size Received : 31.5 gram

Total Amount : 544 units

Completed : 12/02/23 Expires: 12/02/24

Sample Method : SOP.T.20.010

Page 2 of 5



Terpenes

TESTED

Terpenes	LOD (%)	mg/unit	%	Result (%)	Terpenes	LOD (%)	mg/unit	%	Result (%)
TOTAL TERPENES	0.007	113.37	3.239		HEXAHYDROTHYMOL	0.007	ND	ND	
ALPHA-TERPINOLENE	0.007	28.81	0.823		ISOBORNEOL	0.007	ND	ND	
ALPHA-PINENE	0.007	18.13	0.518		NEROL	0.007	ND	ND	
BETA-PINENE	0.007	9.35	0.267		PULEGONE	0.007	ND	ND	
BETA-MYRCENE	0.007	9.24	0.264		SABINENE	0.007	ND	ND	
LIMONENE	0.007	6.76	0.193		VALENCENE	0.007	ND	ND	
OCIMENE	0.007	3.12	0.089		ALPHA-CEDRENE	0.007	ND	ND	
LINALOOL	0.007	3.01	0.086		CIS-NEROLIDOL	0.007	ND	ND	
BETA-CARYOPHYLLENE	0.007	2.98	0.085						
ALPHA-PHELLANDRENE	0.007	2.66	0.076		Analysis Method : SOP.T.30.061A.FL, SOP.T.40.061A.FL	Weight:	Extraction date:	Extracted by:	
3-CARENE	0.007	1.96	0.056		2076, 585, 3963	0.8546g	11/30/23 18:46:48	2076	
ALPHA-BISABOLOL	0.007	1.96	0.056		Analysis Batch : DA068876TER				
TOTAL TERPINEOL	0.007	1.72	0.049		Instrument Used : DA-GCMS-004				
GUAJOL	0.007	1.65	0.047		Analysis Date : 11/30/23 18:47:04				
ALPHA-HUMULENE	0.007	1.61	0.046		Dilution : 10				
ALPHA-TERPINENE	0.007	1.47	0.042		Reagent : 121622.26				
FENCHYL ALCOHOL	0.007	1.44	0.041		Consumables : 210414634; MKCN9995; CE0123; R1KB14270				
GAMMA-TERPINENE	0.007	0.88	0.025		Pipette : N/A				
FARNESENE	0.001	0.32	0.009		Terpenoid testing is performed utilizing Gas Chromatography Mass Spectrometry. For all Flower samples, the Total Terpenes % is dry-weight corrected.				
BORNEOL	0.013	<1.40	<0.040						
CAMPHENE	0.007	<0.70	<0.020						
CAMPHOR	0.007	<2.10	<0.060						
CARYOPHYLLENE OXIDE	0.007	<0.70	<0.020						
EUCALYPTOL	0.007	<0.70	<0.020						
FENCHONE	0.007	<1.40	<0.040						
ISOPULEGOL	0.007	<0.70	<0.020						
SABINENE HYDRATE	0.007	<0.70	<0.020						
TRANS-NEROLIDOL	0.007	<0.70	<0.020						
CEDROL	0.007	ND	ND						
GERANIOL	0.007	ND	ND						
GERANYL ACETATE	0.007	ND	ND						

Total (%) 3.239

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino

Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164

Signature
12/02/23



4131 SW 47th AVENUE SUITE 1408
DAVIE, FL, 33314, US
(954) 368-7664

Kaycha Labs

FTH-Bubbly Haze #340 WF 3.5g(1/8oz)

FTH-Bubbly Haze #340

Matrix : Flower

Type: Flower-Cured



Certificate of Analysis

PASSED

FLUENT

82 NE 26th street
Miami, FL, 33137, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA31130006-001

Harvest/Lot ID: 8923 3224 6475 6903

Batch# : 8923 3224 6475

6903

Sampled : 11/30/23

Ordered : 11/30/23

Sample Size Received : 31.5 gram

Total Amount : 544 units

Completed : 12/02/23 Expires: 12/02/24

Sample Method : SOP.T.20.010

Page 3 of 5



Pesticides

PASSED

Pesticide	LOD	Units	Action Level	Pass/Fail	Result	Pesticide	LOD	Units	Action Level	Pass/Fail	Result
TOTAL CONTAMINANT LOAD (PESTICIDES)	0.010	ppm	5	PASS	ND	OXAMYL	0.010	ppm	0.5	PASS	ND
TOTAL DIMETHOMORPH	0.010	ppm	0.2	PASS	ND	PACLOBUTRAZOL	0.010	ppm	0.1	PASS	ND
TOTAL PERMETHRIN	0.010	ppm	0.1	PASS	ND	PHOSMET	0.010	ppm	0.1	PASS	ND
TOTAL PYRETHRINS	0.010	ppm	0.5	PASS	ND	PIPERONYL BUTOXIDE	0.010	ppm	3	PASS	ND
TOTAL SPINETORAM	0.010	ppm	0.2	PASS	ND	PRALLETHRIN	0.010	ppm	0.1	PASS	ND
TOTAL SPINOSAD	0.010	ppm	0.1	PASS	ND	PROPICONAZOLE	0.010	ppm	0.1	PASS	ND
ABAMECTIN B1A	0.010	ppm	0.1	PASS	ND	PROPOXUR	0.010	ppm	0.1	PASS	ND
ACEPHATE	0.010	ppm	0.1	PASS	ND	PYRIDABEN	0.010	ppm	0.2	PASS	ND
ACEQUINOCYL	0.010	ppm	0.1	PASS	ND	SPIROMESIFEN	0.010	ppm	0.1	PASS	ND
ACETAMIPRID	0.010	ppm	0.1	PASS	ND	SPIROTETRAMAT	0.010	ppm	0.1	PASS	ND
ALDICARB	0.010	ppm	0.1	PASS	ND	SPIROXAMINE	0.010	ppm	0.1	PASS	ND
AZOXYSTROBIN	0.010	ppm	0.1	PASS	ND	TEBUCONAZOLE	0.010	ppm	0.1	PASS	ND
BIFENAZATE	0.010	ppm	0.1	PASS	ND	THIACLOPRID	0.010	ppm	0.1	PASS	ND
BIFENTHRIN	0.010	ppm	0.1	PASS	ND	THIAMETHOXAM	0.010	ppm	0.5	PASS	ND
BOSCALID	0.010	ppm	0.1	PASS	ND	TRIFLOXYSTROBIN	0.010	ppm	0.1	PASS	ND
CARBARYL	0.010	ppm	0.5	PASS	ND	PENTACHLORONITROBENZENE (PCNB) *	0.010	PPM	0.15	PASS	ND
CARBOFURAN	0.010	ppm	0.1	PASS	ND	PARATHION-METHYL *	0.010	PPM	0.1	PASS	ND
CHLORANTRANILIPROLE	0.010	ppm	1	PASS	ND	CAPTAN *	0.070	PPM	0.7	PASS	ND
CHLORMEQUAT CHLORIDE	0.010	ppm	1	PASS	ND	CHLORDANE *	0.010	PPM	0.1	PASS	ND
CHLORPYRIFOS	0.010	ppm	0.1	PASS	ND	CHLORFENAPYR *	0.010	PPM	0.1	PASS	ND
CLOFENTEZINE	0.010	ppm	0.2	PASS	ND	CYFLUTHRIN *	0.050	PPM	0.5	PASS	ND
COUMAPHOS	0.010	ppm	0.1	PASS	ND	CYPERMETHRIN *	0.050	PPM	0.5	PASS	ND
DAMINOZIDE	0.010	ppm	0.1	PASS	ND						
DIAZINON	0.010	ppm	0.1	PASS	ND	Analized by: 3379, 585, 3963	Weight: 0.9494g	Extraction date: 11/30/23 17:37:26	Extracted by: 3379		
DICHLORVOS	0.010	ppm	0.1	PASS	ND	Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.101.FL (Gainesville), SOP.T.40.102.FL (Davie)					
DIMETHOATE	0.010	ppm	0.1	PASS	ND	Analytical Batch : DA066882PES		Reviewed On : 12/02/23 13:04:19			
ETHOPROPHOS	0.010	ppm	0.1	PASS	ND	Instrument Used : DA-LCMS-003 (PES)		Batch Date : 11/30/23 12:23:55			
ETOFENPROX	0.010	ppm	0.1	PASS	ND	Analyzed Date : 11/30/23 17:49:27					
ETOXAZOLE	0.010	ppm	0.1	PASS	ND	Dilution : 250					
FENHEXAMID	0.010	ppm	0.1	PASS	ND	Reagent : 112823.R13; 040423.08; 112723.R01; 112923.R04; 112823.R03; 112123.R13; 112923.R05					
FENOXYCARB	0.010	ppm	0.1	PASS	ND	Consumables : 326250IW					
FENPYROXIMATE	0.010	ppm	0.1	PASS	ND	Pipette : DA-093; DA-094; DA-219					
FIPRONIL	0.010	ppm	0.1	PASS	ND	Testing for agricultural agents is performed utilizing Liquid Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
FLONICAMID	0.010	ppm	0.1	PASS	ND	Analized by: 450, 585, 3963	Weight: 0.9494g	Extraction date: 11/30/23 17:37:26	Extracted by: 3379		
FLUDIOXONIL	0.010	ppm	0.1	PASS	ND	Analysis Method : SOP.T.30.151.FL (Gainesville), SOP.T.30.151A.FL (Davie), SOP.T.40.151.FL					
HEXYTHIAZOX	0.010	ppm	0.1	PASS	ND	Analytical Batch : DA066883VOL		Reviewed On : 12/01/23 12:03:47			
IMAZALIL	0.010	ppm	0.1	PASS	ND	Instrument Used : DA-GCMS-010		Batch Date : 11/30/23 12:25:50			
IMIDACLOPRID	0.010	ppm	0.4	PASS	ND	Analyzed Date : 11/30/23 18:22:44					
KRESOXIM-METHYL	0.010	ppm	0.1	PASS	ND	Dilution : 250					
MALATHION	0.010	ppm	0.2	PASS	ND	Reagent : 112823.R13; 040423.08; 112723.R14; 112723.R15					
METALAXYL	0.010	ppm	0.1	PASS	ND	Consumables : 326250IW; 14725401					
METHIOCARB	0.010	ppm	0.1	PASS	ND	Pipette : DA-080; DA-146; DA-218					
METHOMYL	0.010	ppm	0.1	PASS	ND	Testing for agricultural agents is performed utilizing Gas Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
MEVINPHOS	0.010	ppm	0.1	PASS	ND						
MYCLOBUTANIL	0.010	ppm	0.1	PASS	ND						
NALED	0.010	ppm	0.25	PASS	ND						

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino

Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164

Signature
12/02/23



Certificate of Analysis

PASSED

FLUENT

82 NE 26th street
Miami, FL, 33137, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA31130006-001

Harvest/Lot ID: 8923 3224 6475 6903

Batch# : 8923 3224 6475 6903

Sampled : 11/30/23

Ordered : 11/30/23


Sample Size Received : 31.5 gram



Total Amount : 544 units

Completed : 12/02/23 Expires: 12/02/24

Sample Method : SOP.T.20.010

Page 4 of 5

	Microbial	PASSED			
Analyte	LOD	Units	Result	Pass / Fail	Action Level
ASPERGILLUS TERREUS			Not Present	PASS	
ASPERGILLUS NIGER			Not Present	PASS	
ASPERGILLUS FUMIGATUS			Not Present	PASS	
ASPERGILLUS FLAVUS			Not Present	PASS	
SALMONELLA SPECIFIC GENE			Not Present	PASS	
ECOLI SHIGELLA			Not Present	PASS	
TOTAL YEAST AND MOLD	10	CFU/g	73000	PASS	100000
Analyzed by: 3336, 585, 3963	Weight: 1.0487g	Extraction date: 11/30/23 17:42:18	Extracted by: 3336,3390	Reviewed On : 12/02/23 11:03:18 Batch Date : 11/30/23 10:23:03	
Analysis Method : SOP.T.40.056C, SOP.T.40.058.FL, SOP.T.40.209.FL			Analytical Batch : DA066873MIC		
Instrument Used : PathogenDx Scanner DA-111,Applied Biosystems Thermocycler DA-010,fisherbrand Isotemp Heat Block DA-020,fisherbrand Isotemp Heat Block DA-049,Fisher Scientific Isotemp Heat Block DA-021					
Analyzed Date : 11/30/23 15:01:21					
Dilution : N/A					
Reagent : 101123.08; 101123.11; 112423.R01; 081023.07; 091523.41					
Consumables : 7568001007					
Pipette : N/A					
Analyzed by: 3336, 585, 3963	Weight: 1.0487g	Extraction date: 11/30/23 17:42:18	Extracted by: 3336,3390	Reviewed On : 12/02/23 15:14:42 Batch Date : 11/30/23 14:04:19	
Analysis Method : SOP.T.40.208 (Gainesville), SOP.T.40.209.FL			Analytical Batch : DA066902TYM		
Instrument Used : Incubator (25-27C) DA-096					
Analyzed Date : 11/30/23 15:00:31					
Dilution : N/A					
Reagent : 101123.08; 101123.11; 112423.R01; 112423.R02					
Consumables : N/A					
Pipette : N/A					
Total yeast and mold testing is performed utilizing MPN and traditional culture based techniques in accordance with F.S. Rule 64ER20-39.					

	Mycotoxins	PASSED					
Analyte	LOD	Units	Result	Pass / Fail	Action Level		
AFLATOXIN B2	0.002	ppm	ND	PASS	0.02		
AFLATOXIN B1	0.002	ppm	ND	PASS	0.02		
OCHRATOXIN A	0.002	ppm	ND	PASS	0.02		
AFLATOXIN G1	0.002	ppm	ND	PASS	0.02		
AFLATOXIN G2	0.002	ppm	ND	PASS	0.02		
Analyzed by: 3379, 585, 3963	Weight: 0.9494g	Extraction date: 11/30/23 17:37:26	Extracted by: 3379	Reviewed On : 12/02/23 13:03:21 Batch Date : 11/30/23 15:48:33			
Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.40.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.102.FL (Davie)							
Analytical Batch : DA066906MYC			Reviewed On : 12/02/23 13:03:21				
Instrument Used : N/A			Batch Date : 11/30/23 15:48:33				
Analyzed Date : 11/30/23 17:50:02							
Dilution : 250							
Reagent : 112823.R13; 040423.08; 112723.R01; 112923.R04; 112823.R03; 112123.R13; 112923.R05							
Consumables : 326250IW							
Pipette : DA-093; DA-094; DA-219							
Mycotoxins testing utilizing Liquid Chromatography with Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.							
	Heavy Metals	PASSED					
Metal	LOD	Units	Result	Pass / Fail	Action Level		
TOTAL CONTAMINANT LOAD METALS							
ARSENIC	0.080	ppm	ND	PASS	1.1		
CADMIUM	0.020	ppm	ND	PASS	0.2		
MERCURY	0.020	ppm	ND	PASS	0.2		
LEAD	0.020	ppm	ND	PASS	0.5		
Analyzed by: 1022, 585, 3963	Weight: 0.2696g	Extraction date: 11/30/23 12:50:05	Extracted by: 1022	Reviewed On : 12/01/23 10:43:03 Batch Date : 11/30/23 09:47:30			
Analysis Method : SOP.T.30.082.FL, SOP.T.40.082.FL							
Analytical Batch : DA066866HEA			Reviewed On : 12/01/23 10:43:03				
Instrument Used : DA-ICPMS-004			Batch Date : 11/30/23 09:47:30				
Analyzed Date : 11/30/23 17:30:55							
Dilution : 50							
Reagent : 102723.R12; 112723.R04; 111623.R11; 112723.R02; 112723.R03; 112023.R22; 110123.49; 111023.R06							
Consumables : 179436; 210508058; 12594-247CD-247C							
Pipette : DA-061; DA-191; DA-216							
Heavy Metals analysis is performed using Inductively Coupled Plasma Mass Spectrometry in accordance with F.S. Rule 64ER20-39.							



4131 SW 47th AVENUE SUITE 1408
DAVIE, FL, 33314, US
(954) 368-7664

Kaycha Labs

FTH-Bubbly Haze #340 WF 3.5g(1/8oz)
FTH-Bubbly Haze #340
Matrix : Flower
Type: Flower-Cured



Certificate of Analysis

PASSED

FLUENT

82 NE 26th street
Miami, FL, 33137, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA31130006-001

Harvest/Lot ID: 8923 3224 6475 6903

Batch# : 8923 3224 6475
6903

Sampled : 11/30/23

Ordered : 11/30/23

Sample Size Received : 31.5 gram

Total Amount : 544 units

Completed : 12/02/23 Expires: 12/02/24

Sample Method : SOP.T.20.010

Page 5 of 5



Filth/Foreign
Material

PASSED



Moisture

PASSED

Analyte	LOD	Units	Result	P/F	Action Level	Analyte	LOD	Units	Result	P/F	Action Level
Filth and Foreign Material	0.100	%	ND	PASS	1	Moisture Content	1.00	%	14.42	PASS	15
Analyzed by: 1879, 3963	Weight: NA	Extraction date: N/A	Extracted by: N/A			Analyzed by: 4056, 585, 3963	Weight: 0.514g	Extraction date: 11/30/23 19:10:36	Extracted by: 4056		
Analysis Method : SOP.T.40.090 Analytical Batch : DA066910FIL Instrument Used : Filth/Foreign Material Microscope Analyzed Date : 11/30/23 21:40:06						Analysis Method : SOP.T.40.021 Analytical Batch : DA066894MOI Instrument Used : DA-003 Moisture Analyzer Analyzed Date : 11/30/23 19:02:49					
Dilution : N/A Reagent : N/A Consumables : N/A Pipette : N/A						Dilution : N/A Reagent : 031523.19; 020123.02 Consumables : N/A Pipette : DA-066					

Filth and foreign material inspection is performed by visual inspection utilizing naked eye and microscope technologies in accordance with F.S. Rule 64ER20-39.

Moisture Content analysis utilizing loss-on-drying technology in accordance with F.S. Rule 64ER20-39.



Water Activity

PASSED

Analyte	LOD	Units	Result	P/F	Action Level
Water Activity	0.010	aw	0.573	PASS	0.65
Analyzed by: 4056, 585, 3963	Weight: 0.903g	Extraction date: 11/30/23 18:50:42	Extracted by: 4056		
Analysis Method : SOP.T.40.019 Analytical Batch : DA066895WAT Instrument Used : DA-028 Rotronic HygroPalm Analyzed Date : 11/30/23 18:45:37					
Dilution : N/A Reagent : 113021.09 Consumables : PS-14 Pipette : N/A					

Water Activity is performed using a Rotronic HygroPalm HP 23-AW in accordance with F.S. Rule 64ER20-39.

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino

Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164

Signature
12/02/23