



Certificate of Analysis

COMPLIANCE FOR RETAIL











Sample: DA31228003-004
 Harvest/Lot ID: 2131 4662 5448 2338
 Batch#: 2131 4662 5448 2338
 Cultivation Facility: Tampa Cultivation
 Processing Facility : Tampa Processing
 Source Facility : Tampa Cultivation
 Seed to Sale# 6411 0990 3974 7480
 Batch Date: 10/02/23
 Sample Size Received: 15.5 gram
 Total Amount: 1867 units
 Retail Product Size: 0.5 gram
 Ordered: 12/27/23
 Sampled: 12/28/23
 Completed: 12/30/23
 Sampling Method: SOP.T.20.010

Dec 30, 2023 | FLUENT
 82 NE 26th street
 Miami, FL, 33137, US




PASSED

Pages 1 of 6

PRODUCT IMAGE	SAFETY RESULTS								MISC.
	 Pesticides PASSED	 Heavy Metals PASSED	 Microbials PASSED	 Mycotoxins PASSED	 Residuals Solvents PASSED	 Filtration PASSED	 Water Activity PASSED	 Moisture NOT TESTED	 Terpenes TESTED

Cannabinoid **PASSED**



	D9-THC	THCA	CBD	CBDA	D8-THC	CBG	CBGA	CBN	THCV	CBDV	CBC
%	86.081	0.059	0.238	ND	0.946	0.800	ND	0.882	0.620	ND	0.686
mg/unit	430.41	0.30	1.19	ND	4.73	4.00	ND	4.41	3.10	ND	3.43
LOD	0.001	0.001		0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
	%	%	%	%	%	%	%	%	%	%	%

Analized by: 3335, 1665, 585, 4044 Weight: 0.1073g Extraction date: 12/28/23 14:36:09 Extracted by: 3335

Analysis Method : SOP.T.40.031, SOP.T.30.031 Reviewed On : 12/29/23 11:01:27
 Analytical Batch : DA067796POT Batch Date : 12/28/23 09:57:08
 Instrument Used : DA-LC-001
 Analyzed Date : 12/28/23 14:36:15

Dilution : 400
 Reagent : 121923.R15; 060723.24; 121923.R12
 Consumables : 947.109; CE0123; 12594-247CD-247C; R1KB14270
 Pipette : DA-079; DA-108; DA-078

Full Spectrum cannabinoid analysis utilizing High Performance Liquid Chromatography with UV detection in accordance with F.S. Rule 64ER20-39.

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino
 Lab Director
 State License # CMTL-0002
 ISO 17025 Accreditation # ISO/IEC
 17025:2017 Accreditation PJLA-
 Testing 97164


 Signature
 12/30/23



Certificate of Analysis

PASSED

FLUENT

82 NE 26th street
Miami, FL, 33137, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA31228003-004
Harvest/Lot ID: 2131 4662 5448 2338

Batch# : 2131 4662 5448 Sample Size Received : 15.5 gram
2338 Total Amount : 1867 units
Sampled : 12/28/23 Completed : 12/30/23 Expires: 12/30/24
Ordered : 12/28/23 Sample Method : SOP.T.20.010

Page 2 of 6

Terpenes				TESTED			
Terpenes	LOD (%)	mg/unit %	Result (%)	Terpenes	LOD (%)	mg/unit %	Result (%)
TOTAL TERPENES	0.007	12.01	2.402	VALENCENE	0.007	ND	ND
LIMONENE	0.007	5.23	1.046	ALPHA-BISABOLOL	0.007	ND	ND
LINALOOL	0.007	1.57	0.313	ALPHA-CEDRENE	0.007	ND	ND
BETA-MYRCENE	0.007	1.51	0.301	ALPHA-PHELLANDRENE	0.007	ND	ND
BETA-CARYOPHYLLENE	0.007	0.92	0.184	ALPHA-TERPINENE	0.007	ND	ND
BETA-PINENE	0.007	0.61	0.121	CIS-NEROLIDOL	0.007	ND	ND
FENCHYL ALCOHOL	0.007	0.52	0.103	GAMMA-TERPINENE	0.007	ND	ND
ALPHA-TERPINOLENE	0.007	0.45	0.090	TRANS-NEROLIDOL	0.007	ND	ND
ALPHA-PINENE	0.007	0.45	0.089	Analyzed by: 2076, 585, 4044 Weight: 1.0308g Extraction date: 12/28/23 15:32:48 Extracted by: 2076 Analysis Method : SOP.T.30.061A.FL, SOP.T.40.061A.FL Analytical Batch : DA067798TER Instrument Used : DA-GCMS-008 Reviewed On : 12/30/23 11:54:02 Analyzed Date : 12/28/23 15:34:21 Batch Date : 12/28/23 09:59:47 Dilution : 10 Reagent : 121622.26 Consumables : 210414634; MKCN9995; CE0123; R1KB14270 Pipette : N/A Terpenoid testing is performed utilizing Gas Chromatography Mass Spectrometry. For all Flower samples, the Total Terpenes % is dry-weight corrected.			
ALPHA-HUMULENE	0.007	0.28	0.056				
OCIMENE	0.007	0.23	0.045				
FARNESENE	0.001	0.16	0.031				
TOTAL TERPINEOL	0.007	0.12	0.023				
3-CARENE	0.007	ND	ND				
BORNEOL	0.013	ND	ND				
CAMPHENE	0.007	ND	ND				
CAMPHOR	0.007	ND	ND				
CARYOPHYLLENE OXIDE	0.007	ND	ND				
CEDROL	0.007	ND	ND				
EUCALYPTOL	0.007	ND	ND				
FENCHONE	0.007	ND	ND				
GERANIOL	0.007	ND	ND				
GERANYL ACETATE	0.007	ND	ND				
GUAIOL	0.007	ND	ND				
HEXAHYDROTHYMOL	0.007	ND	ND				
ISOBORNEOL	0.007	ND	ND				
ISOPULEGOL	0.007	ND	ND				
NEROL	0.007	ND	ND				
PULEGONE	0.007	ND	ND				
SABINENE	0.007	ND	ND				
SABINENE HYDRATE	0.007	ND	ND				
Total (%)			2.402				

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino
Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJA-
Testing 97164

Signature
12/30/23



Certificate of Analysis

PASSED

FLUENT

82 NE 26th street
Miami, FL, 33137, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA31228003-004

Harvest/Lot ID: 2131 4662 5448 2338

Batch# : 2131 4662 5448

2338

Sampled : 12/28/23

Ordered : 12/28/23


Sample Size Received : 15.5 gram

Total Amount : 1867 units

Completed : 12/30/23 Expires: 12/30/24

Sample Method : SOP.T.20.010

Page 3 of 6



Pesticides

PASSED

Pesticide	LOD	Units	Action Level	Pass/Fail	Result	Pesticide	LOD	Units	Action Level	Pass/Fail	Result
TOTAL CONTAMINANT LOAD (PESTICIDES)	0.010	ppm	5	PASS	ND	OXAMYL	0.010	ppm	0.5	PASS	ND
TOTAL DIMETHOMORPH	0.010	ppm	0.2	PASS	ND	PACLOBUTRAZOL	0.010	ppm	0.1	PASS	ND
TOTAL PERMETHRIN	0.010	ppm	0.1	PASS	ND	PHOSMET	0.010	ppm	0.1	PASS	ND
TOTAL PYRETHRINS	0.010	ppm	0.5	PASS	ND	PIPERONYL BUTOXIDE	0.010	ppm	3	PASS	ND
TOTAL SPINETORAM	0.010	ppm	0.2	PASS	ND	PRALLETHRIN	0.010	ppm	0.1	PASS	ND
TOTAL SPINOSAD	0.010	ppm	0.1	PASS	ND	PROPICONAZOLE	0.010	ppm	0.1	PASS	ND
ABAMECTIN B1A	0.010	ppm	0.1	PASS	ND	PROPOXUR	0.010	ppm	0.1	PASS	ND
ACEPHATE	0.010	ppm	0.1	PASS	ND	PYRIDABEN	0.010	ppm	0.2	PASS	ND
ACEQUINOCYL	0.010	ppm	0.1	PASS	ND	SPIROMESIFEN	0.010	ppm	0.1	PASS	ND
ACETAMIPRID	0.010	ppm	0.1	PASS	ND	SPIROTETRAMAT	0.010	ppm	0.1	PASS	ND
ALDICARB	0.010	ppm	0.1	PASS	ND	SPIROXAMINE	0.010	ppm	0.1	PASS	ND
AZOXYSTROBIN	0.010	ppm	0.1	PASS	ND	TEBUCONAZOLE	0.010	ppm	0.1	PASS	ND
BIFENAZATE	0.010	ppm	0.1	PASS	ND	THIACLOPRID	0.010	ppm	0.1	PASS	ND
BIFENTHRIN	0.010	ppm	0.1	PASS	ND	THIAMETHOXAM	0.010	ppm	0.5	PASS	ND
BOSCALID	0.010	ppm	0.1	PASS	ND	TRIFLOXYSTROBIN	0.010	ppm	0.1	PASS	ND
CARBARYL	0.010	ppm	0.5	PASS	ND	PENTACHLORONITROBENZENE (PCNB) *	0.010	PPM	0.15	PASS	ND
CARBOFURAN	0.010	ppm	0.1	PASS	ND	PARATHION-METHYL *	0.010	PPM	0.1	PASS	ND
CHLORANTRILIPROLE	0.010	ppm	1	PASS	ND	CAPTAN *	0.070	PPM	0.7	PASS	ND
CHLORMEQUAT CHLORIDE	0.010	ppm	1	PASS	ND	CHLORDANE *	0.010	PPM	0.1	PASS	ND
CHLORPYRIFOS	0.010	ppm	0.1	PASS	ND	CHLORFENAPYR *	0.010	PPM	0.1	PASS	ND
CLOFENTEZINE	0.010	ppm	0.2	PASS	ND	CYFLUTHRIN *	0.050	PPM	0.5	PASS	ND
COUMAPHOS	0.010	ppm	0.1	PASS	ND	CYPERMETHRIN *	0.050	PPM	0.5	PASS	ND
DAMINOZIDE	0.010	ppm	0.1	PASS	ND						
DIAZINON	0.010	ppm	0.1	PASS	ND	Analyzed by: 3379, 585, 4044 Weight: 0.2453g Extraction date: 12/28/23 16:25:27 Extracted by: 450,3379 Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.101.FL (Gainesville), SOP.T.40.102.FL (Davie) Analytical Batch : DA067810PES Reviewed On : 12/29/23 11:37:32 Instrument Used : DA-LCMS-003 (PES) Batch Date : 12/28/23 11:32:13 Analyzed Date : N/A Dilution : 250 Reagent : 122623.R03; 040423.08; 122623.R01; 122723.R30; 122623.R02; 112123.R13; 122723.R01 Consumables : 326250IW Pipette : DA-093; DA-094; DA-219					
DICHLORVOS	0.010	ppm	0.1	PASS	ND	Analyzed by: 450, 585, 4044 Weight: 0.2453g Extraction date: 12/28/23 16:25:27 Extracted by: 450,3379 Analysis Method : SOP.T.30.151.FL (Gainesville), SOP.T.30.151A.FL (Davie), SOP.T.40.151.FL (Gainesville) Analytical Batch : DA067811VOL Reviewed On : 12/29/23 11:35:41 Instrument Used : DA-GCMS-010 Batch Date : 12/28/23 11:34:28 Analyzed Date : 12/28/23 16:37:10 Dilution : 250 Reagent : 122623.R03; 040423.08; 1121423.R01; 112723.R15 Consumables : 326250IW; 14725401 Pipette : DA-080; DA-146; DA-218					
ETHOPROPHOS	0.010	ppm	0.1	PASS	ND	Testing for agricultural agents is performed utilizing Gas Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
ETOFENPROX	0.010	ppm	0.1	PASS	ND						
ETOXAZOLE	0.010	ppm	0.1	PASS	ND						
FENHEXAMID	0.010	ppm	0.1	PASS	ND						
FENOXYCARB	0.010	ppm	0.1	PASS	ND						
FENPYROXIMATE	0.010	ppm	0.1	PASS	ND						
FIPRONIL	0.010	ppm	0.1	PASS	ND						
FLONICAMID	0.010	ppm	0.1	PASS	ND						
FLUDIOXONIL	0.010	ppm	0.1	PASS	ND						
HEXYTHIAZOX	0.010	ppm	0.1	PASS	ND						
IMAZALIL	0.010	ppm	0.1	PASS	ND						
IMIDACLOPRID	0.010	ppm	0.4	PASS	ND						
KRESOXIM-METHYL	0.010	ppm	0.1	PASS	ND						
MALATHION	0.010	ppm	0.2	PASS	ND						
METALAXYL	0.010	ppm	0.1	PASS	ND						
METHIACARB	0.010	ppm	0.1	PASS	ND						
METHOMYL	0.010	ppm	0.1	PASS	ND						
MEVINPHOS	0.010	ppm	0.1	PASS	ND						
MYCLOBUTANIL	0.010	ppm	0.1	PASS	ND						
NALED	0.010	ppm	0.25	PASS	ND						

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino

Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164

Signature
12/30/23



Certificate of Analysis

PASSED
FLUENT

 82 NE 26th street
 Miami, FL, 33137, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

Sample : DA31228003-004

Harvest/Lot ID: 2131 4662 5448 2338

 Batch# : 2131 4662 5448
 2338

Sampled : 12/28/23

Ordered : 12/28/23

Sample Size Received : 15.5 gram

Total Amount : 1867 units

Completed : 12/30/23 Expires: 12/30/24

Sample Method : SOP.T.20.010

Page 4 of 6



Residual Solvents

PASSED

Solvents	LOD	Units	Action Level	Pass/Fail	Result
1,1-DICHLOROETHENE	0.800	ppm	8	PASS	ND
1,2-DICHLOROETHANE	0.200	ppm	2	PASS	ND
ACETONE	75.000	ppm	750	PASS	ND
DICHLOROMETHANE	12.500	ppm	125	PASS	ND
BENZENE	0.100	ppm	1	PASS	ND
2-PROPANOL	50.000	ppm	500	PASS	ND
CHLOROFORM	0.200	ppm	2	PASS	ND
ETHANOL	500.000	ppm	5000	PASS	ND
ETHYL ACETATE	40.000	ppm	400	PASS	ND
BUTANES (N-BUTANE)	500.000	ppm	5000	PASS	ND
ACETONITRILE	6.000	ppm	60	PASS	ND
ETHYL ETHER	50.000	ppm	500	PASS	ND
ETHYLENE OXIDE	0.500	ppm	5	PASS	ND
HEPTANE	500.000	ppm	5000	PASS	ND
METHANOL	25.000	ppm	250	PASS	ND
N-HEXANE	25.000	ppm	250	PASS	ND
PENTANES (N-PENTANE)	75.000	ppm	750	PASS	ND
TOLUENE	15.000	ppm	150	PASS	ND
TOTAL XYLENES	15.000	ppm	150	PASS	ND
PROPANE	500.000	ppm	5000	PASS	ND
TRICHLOROETHYLENE	2.500	ppm	25	PASS	ND

 Analyzed by:
 850, 585, 4044

 Weight:
 0.0257g

 Extraction date:
 12/29/23 12:26:44

 Extracted by:
 850,585

 Analysis Method : SOP.T.40.041.FL
 Analytical Batch : DA06782650L
 Instrument Used : DA-GCMS-002
 Analyzed Date : 12/29/23 12:30:32

 Reviewed On : 12/30/23 17:42:11
 Batch Date : 12/28/23 14:30:39

 Dilution : 1
 Reagent : N/A
 Consumables : R2017.167; G201.167
 Pipette : DA-309 25 uL Syringe 35028

Residual solvents analysis is performed utilizing Gas Chromatography Mass Spectrometry in accordance with with F.S. Rule 64ER20-39.



Certificate of Analysis

PASSED

FLUENT

82 NE 26th street
Miami, FL, 33137, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA31228003-004
Harvest/Lot ID: 2131 4662 5448 2338
Batch# : 2131 4662 5448 2338
Sample Size Received : 15.5 gram
Total Amount : 1867 units
Sampled : 12/28/23
Completed : 12/30/23 Expires: 12/30/24
Ordered : 12/28/23
Sample Method : SOP.T.20.010

Page 5 of 6

	Microbial	PASSED		Mycotoxins	PASSED
---	------------------	---------------	---	-------------------	---------------

Analyte	LOD	Units	Result	Pass / Fail	Action Level
SALMONELLA SPECIFIC GENE			Not Present	PASS	
ECOLI SHIGELLA			Not Present	PASS	
ASPERGILLUS FLAVUS			Not Present	PASS	
ASPERGILLUS FUMIGATUS			Not Present	PASS	
ASPERGILLUS TERREUS			Not Present	PASS	
ASPERGILLUS NIGER			Not Present	PASS	
TOTAL YEAST AND MOLD	10	CFU/g	<10	PASS	100000

Analyzed by: 3336, 3621, 585, 4044
Weight: 0.896g
Extraction date: 12/28/23 13:07:50
Extracted by: 3336
Analysis Method : SOP.T.40.056C, SOP.T.40.058.FL, SOP.T.40.209.FL
Analytical Batch : DA067793MIC
Instrument Used : Incubator (37°C) DA- 188,DA-265 Gene-UP RTPCR,DA-351 GENE-UP RTPCR,Incubator (42°C) DA- 328
Reviewed On : 12/30/23 15:58:28
Batch Date : 12/28/23 09:08:12
Analyzed Date : 12/29/23 16:12:46
Dilution : N/A
Reagent : 103123.R11; 121923.R17
Consumables : 2125220; 2125230
Pipette : N/A

Analyte	LOD	Units	Result	Pass / Fail	Action Level
AFLATOXIN B2	0.002	ppm	ND	PASS	0.02
AFLATOXIN B1	0.002	ppm	ND	PASS	0.02
OCHRATOXIN A	0.002	ppm	ND	PASS	0.02
AFLATOXIN G1	0.002	ppm	ND	PASS	0.02
AFLATOXIN G2	0.002	ppm	ND	PASS	0.02

Analyzed by: 3379, 585, 4044
Weight: 0.2453g
Extraction date: 12/28/23 16:25:27
Extracted by: 450,3379
Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.40.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.102.FL (Davie)
Analytical Batch : DA067815MYC
Instrument Used : N/A
Analyzed Date : N/A
Reviewed On : 12/29/23 11:07:45
Batch Date : 12/28/23 11:40:53
Dilution : 250
Reagent : 122623.R03; 040423.08; 122623.R01; 122723.R30; 122623.R02; 112123.R13; 122723.R01
Consumables : 326250IW
Pipette : DA-093; DA-094; DA-219

Mycotoxins testing utilizing Liquid Chromatography with Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.

Analyte	LOD	Units	Result	Pass / Fail	Action Level
TOTAL YEAST AND MOLD	10	CFU/g	<10	PASS	100000

Analyzed by: 3336, 3621, 585, 4044
Weight: 0.830g
Extraction date: 12/28/23 13:13:42
Extracted by: 3336
Analysis Method : SOP.T.40.208 (Gainesville), SOP.T.40.209.FL
Analytical Batch : DA067825TYM
Instrument Used : Incubator (25-27°C) DA-096
Reviewed On : 12/30/23 15:59:15
Batch Date : 12/28/23 13:12:16
Analyzed Date : 12/28/23 13:59:18
Dilution : N/A
Reagent : 110723.06; 112423.R02
Consumables : N/A
Pipette : N/A

Total yeast and mold testing is performed utilizing MPN and traditional culture based techniques in accordance with F.S. Rule 64ER20-39.

	Heavy Metals	PASSED
---	---------------------	---------------

Metal	LOD	Units	Result	Pass / Fail	Action Level
TOTAL CONTAMINANT LOAD METALS	0.080	ppm	ND	PASS	1.1
ARSENIC	0.020	ppm	ND	PASS	0.2
CADMIUM	0.020	ppm	ND	PASS	0.2
MERCURY	0.020	ppm	ND	PASS	0.2
LEAD	0.020	ppm	ND	PASS	0.5

Analyzed by: 1022, 1879, 585, 4044
Weight: 0.2498g
Extraction date: 12/28/23 15:01:37
Extracted by: 1879,1022

Analysis Method : SOP.T.30.082.FL, SOP.T.40.082.FL
Analytical Batch : DA067808HEA
Instrument Used : DA-ICPMS-004
Analyzed Date : 12/28/23 18:05:21
Reviewed On : 12/29/23 11:52:01
Batch Date : 12/28/23 11:25:01
Dilution : 50
Reagent : 120123.R17; 122623.R06; 121723.R01; 122623.R04; 122623.R05; 122023.R43; 120623.R45
Consumables : 179436; 210508058; 12594-247CD-247C
Pipette : DA-061; DA-191; DA-216

Heavy Metals analysis is performed using Inductively Coupled Plasma Mass Spectrometry in accordance with F.S. Rule 64ER20-39.





Certificate of Analysis

PASSED

FLUENT

82 NE 26th street
Miami, FL, 33137, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA31228003-004
Harvest/Lot ID: 2131 4662 5448 2338
Batch# : 2131 4662 5448 2338
Sample Size Received : 15.5 gram
Total Amount : 1867 units
Sampled : 12/28/23
Completed : 12/30/23 Expires: 12/30/24
Ordered : 12/28/23
Sample Method : SOP.T.20.010

Page 6 of 6

	Filth/Foreign Material	PASSED
---	-------------------------------	---------------

Analyte	LOD	Units	Result	P/F	Action Level
Filth and Foreign Material	0.100	%	ND	PASS	1

Analyzed by: 1879, 585, 4044	Weight: NA	Extraction date: N/A	Extracted by: N/A
---------------------------------	---------------	-------------------------	----------------------

Analysis Method : SOP.T.40.090
Analytical Batch : DA067828FIL
Instrument Used : Filth/Foreign Material Microscope
Analyzed Date : 12/29/23 01:00:01
Reviewed On : 12/29/23 01:09:14
Batch Date : 12/29/23 00:55:09

Dilution : N/A
Reagent : N/A
Consumables : N/A
Pipette : N/A

Filth and foreign material inspection is performed by visual inspection utilizing naked eye and microscope technologies in accordance with F.S. Rule 64ER20-39.

	Water Activity	PASSED
---	-----------------------	---------------

Analyte	LOD	Units	Result	P/F	Action Level
Water Activity	0.010	aw	0.459	PASS	0.85

Analyzed by: 4056, 585, 4044	Weight: 0.347g	Extraction date: 12/28/23 17:07:53	Extracted by: 4056
---------------------------------	-------------------	---------------------------------------	-----------------------

Analysis Method : SOP.T.40.019
Analytical Batch : DA067819WAT
Instrument Used : DA-028 Rotronic HygroPalm
Analyzed Date : 12/28/23 17:01:28
Reviewed On : 12/29/23 11:01:29
Batch Date : 12/28/23 12:03:16

Dilution : N/A
Reagent : 113021.09
Consumables : PS-14
Pipette : N/A

Water Activity is performed using a Rotronic HygroPalm HP 23-AW in accordance with F.S. Rule 64ER20-39.

