



Certificate of Analysis

COMPLIANCE FOR RETAIL

Mar 06, 2024 | FLUENT

5540 W. Executive Drive
Tampa, FL, 33609, US



PASSED

Pages 1 of 6


PRODUCT IMAGE



SAFETY RESULTS



Pesticides
PASSED



Heavy Metals
PASSED



Microbials
PASSED



Mycotoxins
PASSED



Residuals Solvents
PASSED



Filtration
PASSED



Water Activity
PASSED



Moisture
NOT TESTED



Terpenes
TESTED

MISC.



Cannabinoid

PASSED



Total THC

0.193%

Total THC/Container : 57.90 mg



Total CBD

2.091%

Total CBD/Container : 627.30 mg



Total Cannabinoids

2.481%

Total Cannabinoids/Container : 744.30 mg

	D9-THC	THCA	CBD	CBDA	D8-THC	CBG	CBGA	CBN	THCV	CBDV	CBC
%	0.186	0.009	1.881	0.240	ND	0.045	0.014	0.005	ND	0.013	0.088
mg/unit	55.80	2.70	564.30	72.00	ND	13.50	4.20	1.50	ND	3.90	26.40
LOD	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
	%	%	%	%	%	%	%	%	%	%	%

Analyzed by:
3335, 1665, 53, 4044

Weight:
3.0689g

Extraction date:
03/04/24 10:44:04

Extracted by:
1665,3335

Analysis Method : SOP.T.40.031, SOP.T.30.031

Analytical Batch : DA070089POT

Instrument Used : DA-LC-007

Analyzed Date : 03/04/24 11:25:34

Reviewed On : 03/05/24 20:28:19

Batch Date : 03/04/24 07:48:06

Dilution : 400

Reagent : 022124.R04; 030923.08; 020724.R04

Consumables : 947.109; 280670723; CE0123; R1KB14270

Pipette : DA-079; DA-108; DA-078

Full Spectrum cannabinoid analysis utilizing High Performance Liquid Chromatography with UV detection in accordance with F.S. Rule 64ER20-39.

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino

Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164

Signature
03/06/24



4131 SW 47th AVENUE SUITE 1408
DAVIE, FL, 33314, US
(954) 368-7664

Kaycha Labs

Perla Drops 30g
Perla Drops
Matrix : Derivative
usable products)



Type: Products for oral administration (pills, capsules, tinctures, and similar

Certificate of Analysis

PASSED

FLUENT

5540 W. Executive Drive
Tampa, FL, 33609, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA40302013-002

Harvest/Lot ID: 9675 7823 3365 5002

Batch# : 9675 7823 3365
5002

Sampled : 03/02/24
Ordered : 03/02/24

Sample Size Received : 180 gram

Total Amount : 1275 units

Completed : 03/06/24 Expires: 03/06/25

Sample Method : SOP.T.20.010

Page 2 of 6



Terpenes

TESTED

Terpenes	LOD (%)	mg/unit	%	Result (%)	Terpenes	LOD (%)	mg/unit	%	Result (%)
TOTAL TERPENES	0.007	12.00	0.040		ALPHA-PINENE	0.007	ND	ND	
BETA-CARYOPHYLLENE	0.007	8.10	0.027		ALPHA-TERPINENE	0.007	ND	ND	
FARNESENE	0.001	3.90	0.013		ALPHA-TERPINOLENE	0.007	ND	ND	
3-CARENE	0.007	ND	ND		BETA-MYRCENE	0.007	ND	ND	
BORNEOL	0.013	ND	ND		BETA-PINENE	0.007	ND	ND	
CAMPHENE	0.007	ND	ND		CIS-NEROLIDOL	0.007	ND	ND	
CAMPHOR	0.007	ND	ND		GAMMA-TERPINENE	0.007	ND	ND	
CARYOPHYLLENE OXIDE	0.007	ND	ND		TRANS-NEROLIDOL	0.007	ND	ND	
CEDROL	0.007	ND	ND		Analysis by:	Weight:	Extraction date:	Extracted by:	
EUCALYPTOL	0.007	ND	ND		795, 53, 1665, 4044	0.6529g	03/05/24 15:26:20	795	
FENCHONE	0.007	ND	ND		Analysis Method : SOP.T.30.061A.FL, SOP.T.40.061A.FL				
FENCHYL ALCOHOL	0.007	ND	ND		Analytical Batch : DA070096TER			Reviewed On : 03/05/24 17:36:31	
GERANIOL	0.007	ND	ND		Instrument Used : DA-GCMS-004			Batch Date : 03/04/24 14:41:59	
GERANYL ACETATE	0.007	ND	ND		Analyzed Date : N/A				
GUAJOL	0.007	ND	ND		Dilution : 10				
HEXAHYDROTHYMOL	0.007	ND	ND		Reagent : N/A				
ISOBORNEOL	0.007	ND	ND		Consumables : N/A				
ISOPULEGOL	0.007	ND	ND		Pipette : N/A				
LIMONENE	0.007	ND	ND		Terpenoid testing is performed utilizing Gas Chromatography Mass Spectrometry. For all Flower samples, the Total Terpenes % is dry-weight corrected.				
LINALOL	0.007	ND	ND						
NEROL	0.007	ND	ND						
OCIMENE	0.007	ND	ND						
PULEGONE	0.007	ND	ND						
SABINENE	0.007	ND	ND						
SABINENE HYDRATE	0.007	ND	ND						
TOTAL TERPINEOL	0.007	ND	ND						
VALENCENE	0.007	ND	ND						
ALPHA-BISABOLOL	0.007	ND	ND						
ALPHA-CEDRENE	0.007	ND	ND						
ALPHA-HUMULENE	0.007	ND	ND						
ALPHA-PHELLANDRENE	0.007	ND	ND						
Total (%)			0.040						

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino
Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164

Signature
03/06/24



4131 SW 47th AVENUE SUITE 1408
DAVIE, FL, 33314, US
(954) 368-7664

Kaycha Labs

Perla Drops 30g
Perla Drops
Matrix : Derivative
usable products)



Type: Products for oral administration (pills, capsules, tinctures, and similar

Certificate of Analysis

PASSED

FLUENT

5540 W. Executive Drive
Tampa, FL, 33609, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA40302013-002

Harvest/Lot ID: 9675 7823 3365 5002

Batch #: 9675 7823 3365
5002

Sampled : 03/02/24

Ordered : 03/02/24

Sample Size Received : 180 gram

Total Amount : 1275 units

Completed : 03/06/24 Expires: 03/06/25

Sample Method : SOP.T.20.010

Page 3 of 6



Pesticides

PASSED

Pesticide	LOD	Units	Action Level	Pass/Fail	Result	Pesticide	LOD	Units	Action Level	Pass/Fail	Result
TOTAL CONTAMINANT LOAD (PESTICIDES)	0.010	ppm	30	PASS	0.276	OXAMYL	0.010	ppm	0.5	PASS	ND
TOTAL DIMETHOMORPH	0.010	ppm	3	PASS	ND	PACLOBUTRAZOL	0.010	ppm	0.1	PASS	ND
TOTAL PERMETHRIN	0.010	ppm	1	PASS	ND	PHOSMET	0.010	ppm	0.2	PASS	ND
TOTAL PYRETHRINS	0.010	ppm	1	PASS	0.276	PIPERONYL BUTOXIDE	0.010	ppm	3	PASS	ND
TOTAL SPINETORAM	0.010	ppm	3	PASS	ND	PRALLETHRIN	0.010	ppm	0.4	PASS	ND
TOTAL SPINOSAD	0.010	ppm	3	PASS	ND	PROPICONAZOLE	0.010	ppm	1	PASS	ND
ABAMECTIN B1A	0.010	ppm	0.3	PASS	ND	PROPOXUR	0.010	ppm	0.1	PASS	ND
ACEPHATE	0.010	ppm	3	PASS	ND	PYRIDABEN	0.010	ppm	3	PASS	ND
ACEQUINOCYL	0.010	ppm	2	PASS	ND	SPIROMESIFEN	0.010	ppm	3	PASS	ND
ACETAMIPRID	0.010	ppm	3	PASS	ND	SPIROTETRAMAT	0.010	ppm	3	PASS	ND
ALDICARB	0.010	ppm	0.1	PASS	ND	SPIROXAMINE	0.010	ppm	0.1	PASS	ND
AZOXYSTROBIN	0.010	ppm	3	PASS	ND	TEBUCONAZOLE	0.010	ppm	1	PASS	ND
BIFENAZATE	0.010	ppm	3	PASS	ND	THIACLOPRID	0.010	ppm	0.1	PASS	ND
BIFENTHRIN	0.010	ppm	0.5	PASS	ND	THIAMETHOXAM	0.010	ppm	1	PASS	ND
BOSCALID	0.010	ppm	3	PASS	ND	TRIFLOXYSTROBIN	0.010	ppm	3	PASS	ND
CARBARYL	0.010	ppm	0.5	PASS	ND	PENTACHLORONITROBENZENE (PCNB) *	0.010	PPM	0.2	PASS	ND
CARBOFURAN	0.010	ppm	0.1	PASS	ND	PARATHION-METHYL *	0.010	PPM	0.1	PASS	ND
CHLORANTRANILIPROLE	0.010	ppm	3	PASS	ND	CAPTAN *	0.070	PPM	3	PASS	ND
CHLORMEQUAT CHLORIDE	0.010	ppm	3	PASS	ND	CHLORDANE *	0.010	PPM	0.1	PASS	ND
CHLORPYRIFOS	0.010	ppm	0.1	PASS	ND	CHLORFENAPYR *	0.010	PPM	0.1	PASS	ND
CLOFENTEZINE	0.010	ppm	0.5	PASS	ND	CYFLUTHRIN *	0.050	PPM	1	PASS	ND
COUMAPHOS	0.010	ppm	0.1	PASS	ND	CYPERMETHRIN *	0.050	PPM	1	PASS	ND
DAMINOZIDE	0.010	ppm	0.1	PASS	ND						
DIAZINON	0.010	ppm	3	PASS	ND	Analysis by:	Weight:	Extraction date:	Extracted by:		
DICHLORVOS	0.010	ppm	0.1	PASS	ND	3379, 1665, 4044	0.2358g	03/04/24 12:41:36	4056,3379		
ETHOPROPHOS	0.010	ppm	0.1	PASS	ND	Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.101.FL (Gainesville),					
ETOFENPROX	0.010	ppm	0.1	PASS	ND	SOP.T.40.102.FL (Davie)					
ETOXAZOLE	0.010	ppm	1.5	PASS	ND	Analytical Batch : DA070062PES			Reviewed On : 03/06/24 18:32:57		
FENHEXAMID	0.010	ppm	3	PASS	ND	Instrument Used : DA-LCMS-003 (PES)			Batch Date : 03/02/24 14:43:15		
FENOXYCARB	0.010	ppm	0.1	PASS	ND	Analyzed Date : 03/04/24 13:19:16					
FENPYROXIMATE	0.010	ppm	2	PASS	ND	Dilution : 250					
FIPRONIL	0.010	ppm	0.1	PASS	ND	Reagent : 022824.R01; 040423.08; 022924.R03; 022824.R04; 022624.R13; 021324.R05; 022824.R02					
FLONICAMID	0.010	ppm	2	PASS	ND	Consumables : 326250IW					
FLUDIOXONIL	0.010	ppm	3	PASS	ND	Pipette : DA-093; DA-094; DA-219					
HEXYTHIAZOX	0.010	ppm	2	PASS	ND	Testing for agricultural agents is performed utilizing Liquid Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
IMAZALIL	0.010	ppm	0.1	PASS	ND	Analysis by:	Weight:	Extraction date:	Extracted by:		
IMIDACLOPRID	0.010	ppm	1	PASS	ND	450, 1665, 4044	0.2358g	03/04/24 12:41:36	4056,3379		
KRESOXIM-METHYL	0.010	ppm	1	PASS	ND	Analysis Method : SOP.T.30.151.FL (Gainesville), SOP.T.30.151A.FL (Davie), SOP.T.40.151.FL					
MALATHION	0.010	ppm	2	PASS	ND	Analytical Batch : DA070076VOL			Reviewed On : 03/06/24 09:21:31		
METALAXYL	0.010	ppm	3	PASS	ND	Instrument Used : DA-GCMS-010			Batch Date : 03/03/24 12:05:28		
METHIOCARB	0.010	ppm	0.1	PASS	ND	Analyzed Date : 03/04/24 14:24:29					
METHOMYL	0.010	ppm	0.1	PASS	ND	Dilution : 250					
MEVINPHOS	0.010	ppm	0.1	PASS	ND	Reagent : 022824.R01; 040423.08; 021424.R18; 021424.R19					
MYCLOBUTANIL	0.010	ppm	3	PASS	ND	Consumables : 326250IW; 14725401					
NALED	0.010	ppm	0.5	PASS	ND	Pipette : DA-080; DA-146; DA-218					

Testing for agricultural agents is performed utilizing Gas Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino

Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164

Signature
03/06/24



Certificate of Analysis

PASSED
FLUENT

 5540 W. Executive Drive
 Tampa, FL, 33609, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

Sample : DA40302013-002

Harvest/Lot ID: 9675 7823 3365 5002

 Batch# : 9675 7823 3365
 5002

 Sampled : 03/02/24
 Ordered : 03/02/24

Sample Size Received : 180 gram

Total Amount : 1275 units

Completed : 03/06/24 Expires: 03/06/25

Sample Method : SOP.T.20.010

Page 4 of 6



Residual Solvents

PASSED

Solvents	LOD	Units	Action Level	Pass/Fail	Result
1,1-DICHLOROETHENE	0.800	ppm	8	PASS	ND
1,2-DICHLOROETHANE	0.200	ppm	2	PASS	ND
ACETONE	75.000	ppm	750	PASS	ND
DICHLOROMETHANE	12.500	ppm	125	PASS	ND
BENZENE	0.100	ppm	1	PASS	ND
2-PROPANOL	50.000	ppm	500	PASS	ND
CHLOROFORM	0.200	ppm	2	PASS	ND
ETHANOL	500.000	ppm		TESTED	ND
ETHYL ACETATE	40.000	ppm	400	PASS	ND
BUTANES (N-BUTANE)	500.000	ppm	5000	PASS	ND
ACETONITRILE	6.000	ppm	60	PASS	ND
ETHYL ETHER	50.000	ppm	500	PASS	ND
ETHYLENE OXIDE	0.500	ppm	5	PASS	ND
HEPTANE	500.000	ppm	5000	PASS	ND
METHANOL	25.000	ppm	250	PASS	ND
N-HEXANE	25.000	ppm	250	PASS	ND
PENTANES (N-PENTANE)	75.000	ppm	750	PASS	ND
TOLUENE	15.000	ppm	150	PASS	ND
TOTAL XYLENES	15.000	ppm	150	PASS	ND
PROPANE	500.000	ppm	5000	PASS	ND
TRICHLOROETHYLENE	2.500	ppm	25	PASS	ND

 Analyzed by:
 850, 53, 1665, 4044

 Weight:
 0.0276g

 Extraction date:
 03/05/24 13:09:52

 Extracted by:
 850

 Analysis Method : SOP.T.40.041.FL
 Analytical Batch : DA070092SOL
 Instrument Used : DA-GCMS-003
 Analyzed Date : 03/04/24 12:38:43

 Reviewed On : 03/05/24 17:44:18
 Batch Date : 03/04/24 11:14:18

 Dilution : 1
 Reagent : N/A
 Consumables : N/A
 Pipette : N/A

Residual solvents analysis is performed utilizing Gas Chromatography Mass Spectrometry in accordance with with F.S. Rule 64ER20-39.



Certificate of Analysis

PASSED
FLUENT

 5540 W. Executive Drive
 Tampa, FL, 33609, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

Sample : DA40302013-002

Harvest/Lot ID: 9675 7823 3365 5002

 Batch# : 9675 7823 3365
 5002

Sampled : 03/02/24

Ordered : 03/02/24




Sample Size Received : 180 gram

Total Amount : 1275 units

Completed : 03/06/24 Expires: 03/06/25

Sample Method : SOP.T.20.010

Page 5 of 6

 Microbial PASSED						 Mycotoxins PASSED					
Analyte	LOD	Units	Result	Pass / Fail	Action Level	Analyte	LOD	Units	Result	Pass / Fail	Action Level
ASPERGILLUS TERREUS			Not Present	PASS		AFLATOXIN B2	0.002	ppm	ND	PASS	0.02
ASPERGILLUS NIGER			Not Present	PASS		AFLATOXIN B1	0.002	ppm	ND	PASS	0.02
ASPERGILLUS FUMIGATUS			Not Present	PASS		OCHRATOXIN A	0.002	ppm	ND	PASS	0.02
ASPERGILLUS FLAVUS			Not Present	PASS		AFLATOXIN G1	0.002	ppm	ND	PASS	0.02
SALMONELLA SPECIFIC GENE			Not Present	PASS		AFLATOXIN G2	0.002	ppm	ND	PASS	0.02
ECOLI SHIGELLA			Not Present	PASS							
TOTAL YEAST AND MOLD	10	CFU/g	<10	PASS	100000						
Analyzed by: 3390, 53, 1665, 4044 Weight: 0.8982g Extraction date: 03/03/24 12:24:00 Extracted by: 4044 Analysis Method : SOP.T.40.056C, SOP.T.40.058.FL, SOP.T.40.209.FL Analytical Batch : DA070072MIC Reviewed On : 03/05/24 19:54:47 Batch Date : 03/03/24 10:16:46 Instrument Used : PathogenDx Scanner DA-111,Applied Biosystems Thermocycler DA-171,fisherbrand Isotemp Heat Block DA-020,fisherbrand Isotemp Heat Block DA-049,Fisher Scientific Isotemp Heat Block DA-021 Analyzed Date : 03/04/24 13:25:30 Dilution : N/A Reagent : 012424.42; 012424.47; 022224.R10; 083123.107 Consumables : 7569001065 Pipette : N/A						Analyzed by: 3379, 1665, 4044 Weight: 0.2358g Extraction date: 03/04/24 12:41:36 Extracted by: 4056,3379 Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.40.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.102.FL (Davie) Analytical Batch : DA070078MYC Instrument Used : N/A Analyzed Date : 03/04/24 13:19:24 Dilution : 250 Reagent : 022824.R01; 040423.08; 022924.R03; 022824.R04; 022624.R13; 021324.R05; 022824.R02 Consumables : 326250IW Pipette : DA-093; DA-094; DA-219 Mycotoxins testing utilizing Liquid Chromatography with Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
 Heavy Metals PASSED											
Metal	LOD	Units	Result	Pass / Fail	Action Level						
TOTAL CONTAMINANT LOAD METALS	0.080	ppm	ND	PASS	5						
ARSENIC	0.020	ppm	ND	PASS	1.5						
CADMIUM	0.020	ppm	ND	PASS	0.5						
MERCURY	0.020	ppm	ND	PASS	3						
LEAD	0.020	ppm	ND	PASS	0.5						
Analyzed by: 1022, 1665, 4044 Weight: 0.2524g Extraction date: 03/04/24 11:12:14 Extracted by: 4306,1022 Analysis Method : SOP.T.30.082.FL, SOP.T.40.082.FL Analytical Batch : DA070079HEA Instrument Used : DA-ICPMS-004 Analyzed Date : 03/04/24 18:06:49 Dilution : 50 Reagent : 020724.R07; 021324.R02; 030424.R04; 030424.R01; 030424.R02; 030424.R03; 030424.01 Consumables : 179436; 34623011; 210508058 Pipette : DA-061; DA-191; DA-216 Heavy Metals analysis is performed using Inductively Coupled Plasma Mass Spectrometry in accordance with F.S. Rule 64ER20-39.											



Certificate of Analysis

PASSED
FLUENT

 5540 W. Executive Drive
 Tampa, FL, 33609, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

Sample : DA40302013-002

Harvest/Lot ID: 9675 7823 3365 5002

 Batch# : 9675 7823 3365
 5002

Sampled : 03/02/24

Ordered : 03/02/24

Sample Size Received : 180 gram

Total Amount : 1275 units

Completed : 03/06/24 Expires: 03/06/25

Sample Method : SOP.T.20.010

Page 6 of 6


**Filth/Foreign
Material**
PASSED

Analyte	LOD	Units	Result	P/F	Action Level
Filth and Foreign Material	0.100	%	ND	PASS	1

Analyzed by:	Weight:	Extraction date:	Extracted by:
1665, 4044	NA	N/A	N/A

Analysis Method : SOP.T.40.090

Analytical Batch : DA070098FIL

Instrument Used : N/A

Analyzed Date : N/A

Reviewed On : 03/05/24 09:20:01

Batch Date : 03/05/24 09:08:10

Dilution : N/A

Reagent : N/A

Consumables : N/A

Pipette : N/A

Filth and foreign material inspection is performed by visual inspection utilizing naked eye and microscope technologies in accordance with F.S. Rule 64ER20-39.


Water Activity
PASSED

Analyte	LOD	Units	Result	P/F	Action Level
Water Activity	0.010	aw	0.427	TESTED	

Analyzed by:	Weight:	Extraction date:	Extracted by:
4444, 53, 1665, 4044	1.15g	03/03/24 14:03:04	4444

Analysis Method : SOP.T.40.019

Analytical Batch : DA070071WAT

Instrument Used : DA256 Rotronic HygroPalm

Analyzed Date : 03/03/24 13:31:52

Reviewed On : 03/05/24 17:47:53

Batch Date : 03/03/24 10:15:54

Dilution : N/A

Reagent : 022024.28

Consumables : PS-14

Pipette : N/A

Water Activity is performed using a Rotronic HygroPalm HP 23-AW in accordance with F.S. Rule 64ER20-39.