



Certificate of Analysis

COMPLIANCE FOR RETAIL

Sample: DA40322002-005
Harvest/Lot ID: 3849 4902 8566 8617
Batch#: 3849 4902 8566 8617
Cultivation Facility: Tampa Cultivation
Processing Facility : Tampa Processing
Source Facility : Tampa Cultivation
Seed to Sale# 9922 8144 9348 3554
Batch Date: 12/29/23
Sample Size Received: 67.5 gram
Total Amount: 1348 units
Retail Product Size: 11.25 gram
Retail Serving Size: 11.25 gram
Servings: 1
Ordered: 03/22/24
Sampled: 03/22/24
Completed: 03/26/24
Sampling Method: SOP.T.20.010

Mar 26, 2024 | FLUENT

5540 W. Executive Drive
 Tampa, FL, 33609, US



PASSED

Pages 1 of 6


PRODUCT IMAGE



SAFETY RESULTS



Pesticides
PASSED



Heavy Metals
PASSED



Microbials
PASSED



Mycotoxins
PASSED



Residuals Solvents
PASSED



Filtration
PASSED



Water Activity
PASSED



Moisture
 NOT TESTED



Terpenes
TESTED

MISC.



Cannabinoid

PASSED



Total THC
4.131%
 Total THC/Container : 464.74 mg



Total CBD
0.013%
 Total CBD/Container : 1.46 mg



Total Cannabinoids
4.412%
 Total Cannabinoids/Container : 496.35 mg

	D9-THC	THCA	CBD	CBDA	D8-THC	CBG	CBGA	CBN	THCV	CBDV	CBC
%	4.127	0.005	0.013	ND	0.016	0.122	0.008	0.034	0.031	ND	0.056
mg/unit	464.29	0.56	1.46	ND	1.80	13.73	0.90	3.83	3.49	ND	6.30
LOD	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001	0.001
	%	%	%	%	%	%	%	%	%	%	%

Analyzed by:
 1665, 585, 1440

Weight:
 2.8137g

Extraction date:
 03/22/24 13:32:16

Extracted by:
 3335

Analysis Method : SOP.T.40.031, SOP.T.30.031
 Analytical Batch : DA070774POT
 Instrument Used : DA-LC-007
 Analyzed Date : 03/22/24 13:57:49

Reviewed On : 03/25/24 11:56:17
 Batch Date : 03/22/24 11:13:01

Dilution : 400
 Reagent : 022724.R01; 060723.24; 030824.R01
 Consumables : 947.109; 280670723; CE0123; R1KB14270
 Pipette : DA-079; DA-108; DA-078

Full Spectrum cannabinoid analysis utilizing High Performance Liquid Chromatography with UV detection in accordance with F.S. Rule 64ER20-39.

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino
 Lab Director

State License # CMTL-0002
 ISO 17025 Accreditation # ISO/IEC
 17025:2017 Accreditation PJLA-
 Testing 97164



Signature
 03/26/24



Type: Products for oral administration (pills, capsules, tinctures, and similar

Certificate of Analysis

PASSED

FLUENT

5540 W. Executive Drive
Tampa, FL, 33609, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA40322002-005

Harvest/Lot ID: 3849 4902 8566 8617

Batch# : 3849 4902 8566
8617

Sampled : 03/22/24

Ordered : 03/22/24

Sample Size Received : 67.5 gram

Total Amount : 1348 units

Completed : 03/26/24 Expires: 03/26/25

Sample Method : SOP.T.20.010

Page 2 of 6

Terpenes				TESTED			
Terpenes	LOD (%)	mg/unit %	Result (%)	Terpenes	LOD (%)	mg/unit %	Result (%)
TOTAL TERPENES	0.007	81.90	0.728	SABINENE	0.007	ND	ND
ALPHA-TERPINOLENE	0.007	31.73	0.282	SABINENE HYDRATE	0.007	ND	ND
BETA-MYRCENE	0.007	9.00	0.080	TOTAL TERPINEOL	0.007	ND	ND
OCIMENE	0.007	6.98	0.062	ALPHA-CEDRENE	0.007	ND	ND
BETA-CARYOPHYLLENE	0.007	5.74	0.051	ALPHA-PHELLANDRENE	0.007	ND	ND
LIMONENE	0.007	5.40	0.048	ALPHA-TERPINENE	0.007	ND	ND
VALENCENE	0.007	3.60	0.032	CIS-NEROLIDOL	0.007	ND	ND
ALPHA-BISABOLOL	0.007	3.49	0.031	TRANS-NEROLIDOL	0.007	ND	ND
ALPHA-HUMULENE	0.007	2.93	0.026	Analysis Method : SOP.T.30.061A.FL, SOP.T.40.061A.FL			
BETA-PINENE	0.007	2.93	0.026	Analyzed by : 3605, 585, 1440	Weight : 0.2139g	Extraction date : 03/22/24 14:39:58	Extracted by : 3605
LINALOOL	0.007	2.81	0.025	Analysis Batch : DA070786TER			
FENCHYL ALCOHOL	0.007	2.70	0.024	Instrument Used : DA-GCMS-004			Reviewed On : 03/25/24 11:56:19
ALPHA-PINENE	0.007	2.36	0.021	Analyzed Date : 03/22/24 14:40:29			Batch Date : 03/22/24 11:38:30
GAMMA-TERPINENE	0.007	2.25	0.020	Dilution : 10			
3-CARENE	0.007	ND	ND	Reagent : 022224.01			
BORNEOL	0.013	ND	ND	Consumables : 947.109; CE0123			
CAMPHENE	0.007	ND	ND	Pipette : DA-063			
CAMPHOR	0.007	ND	ND	Terpenoid testing is performed utilizing Gas Chromatography Mass Spectrometry. For all Flower samples, the Total Terpenes % is dry-weight corrected.			
CARYOPHYLLENE OXIDE	0.007	ND	ND				
CEDROL	0.007	ND	ND				
EUCALYPTOL	0.007	ND	ND				
FARNESENE	0.001	ND	ND				
FENCHONE	0.007	ND	ND				
GERANIOL	0.007	ND	ND				
GERANYL ACETATE	0.007	ND	ND				
GUAJOL	0.007	ND	ND				
HEXAHYDROTHYMOL	0.007	ND	ND				
ISOBORNEOL	0.007	ND	ND				
ISOPULEGOL	0.007	ND	ND				
NEROL	0.007	ND	ND				
PULEGONE	0.007	ND	ND				
Total (%)			0.728				

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino
Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJA-
Testing 97164

Signature
03/26/24



Type: Products for oral administration (pills, capsules, tinctures, and similar

Certificate of Analysis

PASSED

FLUENT

5540 W. Executive Drive
Tampa, FL, 33609, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA40322002-005

Harvest/Lot ID: 3849 4902 8566 8617

Batch# : 3849 4902 8566

8617

Sampled : 03/22/24

Ordered : 03/22/24


Sample Size Received : 67.5 gram

Total Amount : 1348 units

Completed : 03/26/24 Expires: 03/26/25

Sample Method : SOP.T.20.010

Page 3 of 6



Pesticides

PASSED

Pesticide	LOD	Units	Action Level	Pass/Fail	Result	Pesticide	LOD	Units	Action Level	Pass/Fail	Result
TOTAL CONTAMINANT LOAD (PESTICIDES)	0.010	ppm	30	PASS	ND	OXAMYL	0.010	ppm	0.5	PASS	ND
TOTAL DIMETHOMORPH	0.010	ppm	3	PASS	ND	PACLOBUTRAZOL	0.010	ppm	0.1	PASS	ND
TOTAL PERMETHRIN	0.010	ppm	1	PASS	ND	PHOSMET	0.010	ppm	0.2	PASS	ND
TOTAL PYRETHRINS	0.010	ppm	1	PASS	ND	PIPERONYL BUTOXIDE	0.010	ppm	3	PASS	ND
TOTAL SPINETORAM	0.010	ppm	3	PASS	ND	PRALLETHRIN	0.010	ppm	0.4	PASS	ND
TOTAL SPINOSAD	0.010	ppm	3	PASS	ND	PROPICONAZOLE	0.010	ppm	1	PASS	ND
ABAMECTIN B1A	0.010	ppm	0.3	PASS	ND	PROPOXUR	0.010	ppm	0.1	PASS	ND
ACEPHATE	0.010	ppm	3	PASS	ND	PYRIDABEN	0.010	ppm	3	PASS	ND
ACEQUINOCYL	0.010	ppm	2	PASS	ND	SPIROMESIFEN	0.010	ppm	3	PASS	ND
ACETAMIPRID	0.010	ppm	3	PASS	ND	SPIROTETRAMAT	0.010	ppm	3	PASS	ND
ALDICARB	0.010	ppm	0.1	PASS	ND	SPIROXAMINE	0.010	ppm	0.1	PASS	ND
AZOXYSTROBIN	0.010	ppm	3	PASS	ND	TEBUCONAZOLE	0.010	ppm	1	PASS	ND
BIFENAZATE	0.010	ppm	3	PASS	ND	THIACLOPRID	0.010	ppm	0.1	PASS	ND
BIFENTHRIN	0.010	ppm	0.5	PASS	ND	THIAMETHOXAM	0.010	ppm	1	PASS	ND
BOSCALID	0.010	ppm	3	PASS	ND	TRIFLOXYSTROBIN	0.010	ppm	3	PASS	ND
CARBARYL	0.010	ppm	0.5	PASS	ND	PENTACHLORONITROBENZENE (PCNB) *	0.010	PPM	0.2	PASS	ND
CARBOFURAN	0.010	ppm	0.1	PASS	ND	PARATHION-METHYL *	0.010	PPM	0.1	PASS	ND
CHLORANTRANILIPROLE	0.010	ppm	3	PASS	ND	CAPTAN *	0.070	PPM	3	PASS	ND
CHLORMEQUAT CHLORIDE	0.010	ppm	3	PASS	ND	CHLORDANE *	0.010	PPM	0.1	PASS	ND
CHLORPYRIFOS	0.010	ppm	0.1	PASS	ND	CHLORFENAPYR *	0.010	PPM	0.1	PASS	ND
CLOFENTEZINE	0.010	ppm	0.5	PASS	ND	CYFLUTHRIN *	0.050	PPM	1	PASS	ND
COUMAPHOS	0.010	ppm	0.1	PASS	ND	CYPERMETHRIN *	0.050	PPM	1	PASS	ND
DAMINOZIDE	0.010	ppm	0.1	PASS	ND						
DIAZINON	0.010	ppm	3	PASS	ND	Analyzed by: 3379, 585, 1440	Weight: 0.235g	Extraction date: 03/22/24 16:33:13	Extracted by: 3379		
DICHLORVOS	0.010	ppm	0.1	PASS	ND	Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.101.FL (Gainesville), SOP.T.40.102.FL (Davie)					
DIMETHOATE	0.010	ppm	0.1	PASS	ND	Analytical Batch : DA070773PES				Reviewed On : 03/25/24 12:03:20	
ETHOPROPHOS	0.010	ppm	0.1	PASS	ND	Instrument Used : DA-LCMS-003 (PES)				Batch Date : 03/22/24 11:12:26	
ETOFENPROX	0.010	ppm	0.1	PASS	ND	Analyzed Date : 03/22/24 16:42:02					
ETOXAZOLE	0.010	ppm	1.5	PASS	ND	Dilution : 250					
FENHEXAMID	0.010	ppm	3	PASS	ND	Reagent : 031924.R27; 040423.08					
FENOXYCARB	0.010	ppm	0.1	PASS	ND	Consumables : 326250W					
FENPYROXIMATE	0.010	ppm	2	PASS	ND	Pipette : N/A					
FIPRONIL	0.010	ppm	0.1	PASS	ND	Testing for agricultural agents is performed utilizing Liquid Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
FLONICAMID	0.010	ppm	2	PASS	ND	Analyzed by: 450, 585, 1440	Weight: 0.235g	Extraction date: 03/22/24 16:33:13	Extracted by: 3379		
FLUDIOXONIL	0.010	ppm	3	PASS	ND	Analysis Method : SOP.T.30.151.FL (Gainesville), SOP.T.30.151A.FL (Davie), SOP.T.40.151.FL					
HEXYTHIAZOX	0.010	ppm	2	PASS	ND	Analytical Batch : DA070775VOL				Reviewed On : 03/25/24 12:02:12	
IMAZALIL	0.010	ppm	0.1	PASS	ND	Instrument Used : DA-GCMS-001				Batch Date : 03/22/24 11:13:54	
IMIDACLOPRID	0.010	ppm	1	PASS	ND	Analyzed Date : 03/22/24 17:13:16					
KRESOXIM-METHYL	0.010	ppm	1	PASS	ND	Dilution : 250					
MALATHION	0.010	ppm	2	PASS	ND	Reagent : 031924.R27; 040423.08; 031824.R05; 031824.R06					
METALAXYL	0.010	ppm	3	PASS	ND	Consumables : 326250W; 14725401					
METHIACARB	0.010	ppm	0.1	PASS	ND	Pipette : DA-080; DA-146; DA-218					
METHOMYL	0.010	ppm	0.1	PASS	ND	Testing for agricultural agents is performed utilizing Gas Chromatography Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					
MEVINPHOS	0.010	ppm	0.1	PASS	ND						
MYCLOBUTANIL	0.010	ppm	3	PASS	ND						
NALED	0.010	ppm	0.5	PASS	ND						

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino

Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164

Signature
03/26/24



Type: Products for oral administration (pills, capsules, tinctures, and similar

Certificate of Analysis

PASSED
FLUENT

 5540 W. Executive Drive
 Tampa, FL, 33609, US
 Telephone: (305) 900-6266
 Email: Taylor.Jones@getfluent.com

Sample : DA40322002-005

Harvest/Lot ID: 3849 4902 8566 8617

 Batch# : 3849 4902 8566
 8617

Sampled : 03/22/24

Ordered : 03/22/24

Sample Size Received : 67.5 gram

Total Amount : 1348 units

Completed : 03/26/24 Expires: 03/26/25

Sample Method : SOP.T.20.010

Page 4 of 6



Residual Solvents

PASSED

Solvents	LOD	Units	Action Level	Pass/Fail	Result
1,1-DICHLOROETHENE	0.800	ppm	8	PASS	ND
1,2-DICHLOROETHANE	0.200	ppm	2	PASS	ND
ACETONE	75.000	ppm	750	PASS	ND
DICHLOROMETHANE	12.500	ppm	125	PASS	ND
BENZENE	0.100	ppm	1	PASS	ND
2-PROPANOL	50.000	ppm	500	PASS	ND
CHLOROFORM	0.200	ppm	2	PASS	ND
ETHANOL	500.000	ppm		TESTED	ND
ETHYL ACETATE	40.000	ppm	400	PASS	ND
BUTANES (N-BUTANE)	500.000	ppm	5000	PASS	ND
ACETONITRILE	6.000	ppm	60	PASS	ND
ETHYL ETHER	50.000	ppm	500	PASS	ND
ETHYLENE OXIDE	0.500	ppm	5	PASS	ND
HEPTANE	500.000	ppm	5000	PASS	ND
METHANOL	25.000	ppm	250	PASS	ND
N-HEXANE	25.000	ppm	250	PASS	ND
PENTANES (N-PENTANE)	75.000	ppm	750	PASS	ND
TOLUENE	15.000	ppm	150	PASS	ND
TOTAL XYLENES	15.000	ppm	150	PASS	ND
PROPANE	500.000	ppm	5000	PASS	ND
TRICHLOROETHYLENE	2.500	ppm	25	PASS	ND

 Analyzed by:
 850, 585, 1440

 Weight:
 0.0216g

 Extraction date:
 03/24/24 15:23:02

 Extracted by:
 850

 Analysis Method : SOP.T.40.041.FL
 Analytical Batch : DA07079350L
 Instrument Used : DA-GCMS-003
 Analyzed Date : 03/24/24 15:23:26

 Reviewed On : 03/25/24 11:43:29
 Batch Date : 03/22/24 17:01:10

 Dilution : 1
 Reagent : 030420.09
 Consumables : 429651; G201.167
 Pipette : DA-309 25 uL Syringe 35028

Residual solvents analysis is performed utilizing Gas Chromatography Mass Spectrometry in accordance with with F.S. Rule 64ER20-39.



Type: Products for oral administration (pills, capsules, tinctures, and similar

Certificate of Analysis

PASSED

FLUENT

5540 W. Executive Drive
Tampa, FL, 33609, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA40322002-005
Harvest/Lot ID: 3849 4902 8566 8617
Batch# : 3849 4902 8566 Sample Size Received : 67.5 gram
8617 Total Amount : 1348 units
Sampled : 03/22/24 Completed : 03/26/24 Expires: 03/26/25
Ordered : 03/22/24 Sample Method : SOP.T.20.010

Page 5 of 6

	Microbial	PASSED		Mycotoxins	PASSED
---	------------------	---------------	---	-------------------	---------------

Analyte	LOD	Units	Result	Pass / Fail	Action Level
ASPERGILLUS TERREUS			Not Present	PASS	
ASPERGILLUS NIGER			Not Present	PASS	
ASPERGILLUS FUMIGATUS			Not Present	PASS	
ASPERGILLUS FLAVUS			Not Present	PASS	
SALMONELLA SPECIFIC GENE			Not Present	PASS	
ECOLI SHIGELLA			Not Present	PASS	
TOTAL YEAST AND MOLD	10	CFU/g	<10	PASS	100000
Analyzed by: 3390, 585, 1440 Weight: 0.89g Extraction date: 03/22/24 13:18:59 Extracted by: 3390,4044 Analysis Method : SOP.T.40.056C, SOP.T.40.058.FL, SOP.T.40.209.FL Analytical Batch : DA070754MIC Reviewed On : 03/25/24 15:42:55 Instrument Used : PathogenDx Scanner DA-111.fisherbrand Isotemp Heat Block DA-020.fisherbrand Isotemp Heat Block DA-049,Fisher Scientific Isotemp Heat Block DA-021 Analyzed Date : 03/25/24 11:38:18 Batch Date : 03/22/24 10:04:27 Dilution : N/A Reagent : 012424.15; 012424.27; 031824.R18; 091523.42 Consumables : 7569002025; 7569003009 Pipette : N/A					

Analyte	LOD	Units	Result	Pass / Fail	Action Level
AFLATOXIN B2	0.002	ppm	ND	PASS	0.02
AFLATOXIN B1	0.002	ppm	ND	PASS	0.02
OCHRATOXIN A	0.002	ppm	ND	PASS	0.02
AFLATOXIN G1	0.002	ppm	ND	PASS	0.02
AFLATOXIN G2	0.002	ppm	ND	PASS	0.02
Analyzed by: 3379, 585, 1440 Weight: 0.235g Extraction date: 03/22/24 16:33:13 Extracted by: 3379 Analysis Method : SOP.T.30.101.FL (Gainesville), SOP.T.40.101.FL (Gainesville), SOP.T.30.102.FL (Davie), SOP.T.40.102.FL (Davie) Analytical Batch : DA070776MYC Reviewed On : 03/25/24 11:55:27 Instrument Used : N/A Batch Date : 03/22/24 11:15:16 Analyzed Date : 03/22/24 16:42:19 Dilution : 250 Reagent : 031924.R27; 040423.08 Consumables : 326250IW Pipette : N/A Mycotoxins testing utilizing Liquid Chromatography with Triple-Quadrupole Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					

Analyte	LOD	Units	Result	Pass / Fail	Action Level
TOTAL CONTAMINANT LOAD METALS	0.080	ppm	ND	PASS	5
ARSENIC	0.020	ppm	ND	PASS	1.5
CADMIUM	0.020	ppm	ND	PASS	0.5
MERCURY	0.020	ppm	ND	PASS	3
LEAD	0.020	ppm	ND	PASS	0.5
Analyzed by: 4351, 4451, 585, 1440 Weight: 0.89g Extraction date: 03/22/24 13:18:59 Extracted by: 3390,4044 Analysis Method : SOP.T.40.208 (Gainesville), SOP.T.40.209.FL Analytical Batch : DA070784TYM Reviewed On : 03/25/24 11:56:20 Instrument Used : N/A Batch Date : 03/22/24 11:33:09 Analyzed Date : N/A Dilution : N/A Reagent : 012424.15; 012424.27; 031824.R18; 091523.42 Consumables : N/A Pipette : N/A Total yeast and mold testing is performed utilizing MPN and traditional culture based techniques in accordance with F.S. Rule 64ER20-39.					

[Hg]		Heavy Metals		PASSED	
Metal	LOD	Units	Result	Pass / Fail	Action Level
TOTAL CONTAMINANT LOAD METALS	0.080	ppm	ND	PASS	5
ARSENIC	0.020	ppm	ND	PASS	1.5
CADMIUM	0.020	ppm	ND	PASS	0.5
MERCURY	0.020	ppm	ND	PASS	3
LEAD	0.020	ppm	ND	PASS	0.5
Analyzed by: 1022, 585, 1440 Weight: 0.247g Extraction date: 03/22/24 16:05:34 Extracted by: 1022,1879 Analysis Method : SOP.T.30.082.FL, SOP.T.40.082.FL Analytical Batch : DA070761HEA Reviewed On : 03/26/24 07:43:12 Instrument Used : DA-ICPMS-004 Batch Date : 03/22/24 10:56:59 Analyzed Date : 03/25/24 15:40:47 Dilution : 50 Reagent : 030524.R01; 031424.R03; 031124.R04; 031124.R05; 030424.01 Consumables : 179436; 34623011; 210508058 Pipette : DA-061; DA-191; DA-216 Heavy Metals analysis is performed using Inductively Coupled Plasma Mass Spectrometry in accordance with F.S. Rule 64ER20-39.					

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino
Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation P/LA-
Testing 97164



Signature
03/26/24



Type: Products for oral administration (pills, capsules, tinctures, and similar

Certificate of Analysis

PASSED

FLUENT

5540 W. Executive Drive
Tampa, FL, 33609, US
Telephone: (305) 900-6266
Email: Taylor.Jones@getfluent.com

Sample : DA40322002-005
Harvest/Lot ID: 3849 4902 8566 8617
Batch# : 3849 4902 8566 8617
Sample Size Received : 67.5 gram
Total Amount : 1348 units
Sampled : 03/22/24
Completed : 03/26/24 Expires: 03/26/25
Ordered : 03/22/24
Sample Method : SOP.T.20.010

Page 6 of 6

	Filth/Foreign Material	PASSED
---	-------------------------------	---------------

Analyte	LOD	Units	Result	P/F	Action Level
Filth and Foreign Material	0.100	%	ND	PASS	1

Analyzed by: 1879, 585, 1440	Weight: NA	Extraction date: N/A	Extracted by: N/A
---------------------------------	---------------	-------------------------	----------------------

Analysis Method : SOP.T.40.090
Analytical Batch : DA070787FIL
Instrument Used : Filth/Foreign Material Microscope
Analyzed Date : 03/22/24 21:53:51
Reviewed On : 03/22/24 22:39:22
Batch Date : 03/22/24 12:49:10

Dilution : N/A
Reagent : N/A
Consumables : N/A
Pipette : N/A

Filth and foreign material inspection is performed by visual inspection utilizing naked eye and microscope technologies in accordance with F.S. Rule 64ER20-39.

	Water Activity	PASSED
---	-----------------------	---------------

Analyte	LOD	Units	Result	P/F	Action Level
Water Activity	0.010	aw	0.470	TESTED	

Analyzed by: 4056, 585, 1440	Weight: 0.419g	Extraction date: 03/22/24 17:38:32	Extracted by: 4056
---------------------------------	-------------------	---------------------------------------	-----------------------

Analysis Method : SOP.T.40.019
Analytical Batch : DA070790WAT
Instrument Used : DA-028 Rotronic HygroPalm
Analyzed Date : 03/22/24 17:07:44
Reviewed On : 03/25/24 09:55:29
Batch Date : 03/22/24 12:49:52

Dilution : N/A
Reagent : 022024.28
Consumables : PS-14
Pipette : N/A

Water Activity is performed using a Rotronic HygroPalm HP 23-AW in accordance with F.S. Rule 64ER20-39.

This Kaycha Labs Certification shall not be reproduced, unless in its entirety, without written approval from Kaycha Labs. The results relate only to the material or product analyzed. ND=Not Detected, ppm=Parts Per Million, ppb=Parts Per Billion, RSD=Relative Standard Deviation. Limit of Detection (LOD) and Limit Of Quantitation (LOQ) are terms used to describe the smallest concentration that can be detected and reliably measured by an analytical procedure, respectively. Action Levels are State determined thresholds based on F.S. Rule 64ER20-39 and F.S. Rule 5K-4. The Measurement of Uncertainty (MU) error is available from the lab upon request. The "Decision Rule" for pass/fail does not include the MU. Any calculated totals may contain rounding errors.

Vivian Celestino
Lab Director

State License # CMTL-0002
ISO 17025 Accreditation # ISO/IEC
17025:2017 Accreditation PJLA-
Testing 97164



Signature
03/26/24